


LB-136352

7471B INSTRUMENT ID - CV1



Linear

A= 0.0000e+000

B= 2.2140e-004

C= -6.4018e-002

Rho= 0.9999100

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	6/D
0.0	0.000	-0.061	-0.061	13	0.000	13					0
0.2	0.200	0.189	-0.011	1145	0.0 %	1145					1
2.5	2.500	2.546	0.046	11790	0.0 %	11790					-6
5.0	5.000	5.080	0.080	23235	0.0 %	23235					2
7.5	7.500	7.492	-0.008	34130	0.0 %	34130					2
10.0	10.000	9.953	-0.047	45243	0.0 %	45243					0

LB136352 INSTRUMENT ID: CV1

Method: 7471B

Operator: Admin

Date of Analysis: 02 Jul 2025 09:59:20

Type	Sample ID	Extended ID	Conc.	μ Abs.	Units	Type	Std Conc	Method	Date
S	0.0 - 1	50	-	13	PPB	Std	0.0000	7471B	02 Jul 2025 10:09:13
S	0.2 - 1	50.2	-	1145	PPB	Std	0.2000	7471B	02 Jul 2025 10:11:30
S	2.5 - 1	52.5	-	11790	PPB	Std	2.5000	7471B	02 Jul 2025 10:13:47
S	5.0 - 1	55.0	-	23235	PPB	Std	5.0000	7471B	02 Jul 2025 10:16:03
S	7.5 - 1	57.5	-	34130	PPB	Std	7.5000	7471B	02 Jul 2025 10:18:22
S	10.0 - 1	510	-	45243	PPB	Std	10.0000	7471B	02 Jul 2025 10:20:43
U	ICV14 - 1	ICV14	4.1404	18990	PPB	SMPL	-7471B		02 Jul 2025 10:33:37
U	ICB14 - 1	ICB14	-0.0565	34	PPB	SMPL	-7471B		02 Jul 2025 10:38:56
U	CCV44 - 1	CCV44	4.9682	22729	PPB	SMPL	-7471B		02 Jul 2025 10:44:18
U	CCB44 - 1	CCB44	-0.0908	-121	PPB	SMPL	-7471B		02 Jul 2025 10:49:44
U	CRA - 1	CRA	0.1924	1158	PPB	SMPL	-7471B		02 Jul 2025 10:54:36
U	HighStd - 1	HighStd	9.4717	43070	PPB	SMPL	-7471B		02 Jul 2025 11:03:23
U	ChkStd - 1	ChkStd	6.4604	29469	PPB	SMPL	-7471B		02 Jul 2025 11:08:48
U	PB168688BL - 1	PBS	-0.0994	-160	PPB	SMPL	-7471B		02 Jul 2025 11:12:06
U	PB168688BS - 1	LCSS	3.6121	16604	PPB	SMPL	-7471B		02 Jul 2025 11:14:23
U	Q2449-01 - 1	RBR200032	-0.0525	52	PPB	SMPL	-7471B		02 Jul 2025 11:16:40
U	Q2450-01 - 1	PL-2-062725	0.5451	2751	PPB	SMPL	-7471B		02 Jul 2025 11:19:00
U	Q2457-01 - 1	OK-2-062725	0.3389	1820	PPB	SMPL	-7471B		02 Jul 2025 11:21:17
U	Q2458-01 - 1	TP-76	0.1397	920	PPB	SMPL	-7471B		02 Jul 2025 11:23:33
U	Q2458-02 - 1	TP-55	0.1107	789	PPB	SMPL	-7471B		02 Jul 2025 11:25:49
U	CCV45 - 1	CCV45	4.5635	20901	PPB	SMPL	-7471B		02 Jul 2025 11:35:03
U	CCB45 - 1	CCB45	-0.0496	65	PPB	SMPL	-7471B		02 Jul 2025 11:40:23
U	Q2458-03 - 1	TP-68	0.7693	3764	PPB	SMPL	-7471B		02 Jul 2025 11:42:40
U	Q2458-04 - 1	TP-67	0.2240	1301	PPB	SMPL	-7471B		02 Jul 2025 11:44:55
U	Q2458-05 - 1	TP-66	0.2329	1341	PPB	SMPL	-7471B		02 Jul 2025 11:47:13
U	Q2458-06 - 1	TP-60	0.5470	2760	PPB	SMPL	-7471B		02 Jul 2025 11:49:28
U	Q2458-07 - 1	TP-62	0.3646	1936	PPB	SMPL	-7471B		02 Jul 2025 11:51:44
U	Q2458-08 - 1	TP-63	0.5464	2757	PPB	SMPL	-7471B		02 Jul 2025 11:54:01
U	Q2458-09 - 1	TP-59	1.0259	4923	PPB	SMPL	-7471B		02 Jul 2025 11:56:18
U	Q2459-01 - 1	TP-1	1.5347	7221	PPB	SMPL	-7471B		02 Jul 2025 11:58:38
U	Q2464-01 - 1	OR-3-063025	0.3916	2058	PPB	SMPL	-7471B		02 Jul 2025 12:00:58
U	Q2469-01 - 1	WC-1	3.6726	16877	PPB	SMPL	-7471B		02 Jul 2025 12:03:18
U	CCV46 - 1	CCV46	4.8285	22098	PPB	SMPL	-7471B		02 Jul 2025 12:37:11
U	CCB46 - 1	CCB46	-0.0450	86	PPB	SMPL	-7471B		02 Jul 2025 12:42:20
U	Q2469-01DUP - 1	WC-1DUP	3.6484	16768	PPB	SMPL	-7471B		02 Jul 2025 12:47:57
U	Q2469-01MS - 1	WC-1MS	7.4308	33852	PPB	SMPL	-7471B		02 Jul 2025 12:53:29
U	Q2469-01MSD - 1	WC-1MSD	7.4722	34039	PPB	SMPL	-7471B		02 Jul 2025 12:55:44
U	Q2469-01LX5 - 1		0.4627	2379	PPB	SMPL	-7471B		02 Jul 2025 12:58:06
U	Q2469-01A - 1		7.3113	33312	PPB	SMPL	-7471B		02 Jul 2025 13:09:59
U	CCV47 - 1	CCV47	4.9418	22610	PPB	SMPL	-7471B		02 Jul 2025 13:18:38
U	CCB47 - 1	CCB47	-0.0401	108	PPB	SMPL	-7471B		02 Jul 2025 13:23:45