

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 7/3/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:35
In Date: 07/02/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:22
Out Date: 07/03/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB136354

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q2480-01	GPX1	1	1.15	10.52	11.67	10.44	88.3	
Q2480-02	GPX2	2	1.19	10.60	11.79	10.66	89.3	
Q2480-03	GPX3	3	1.15	10.84	11.99	10.77	88.7	
Q2480-04	GPX4	4	1.13	10.84	11.97	10.66	87.9	
Q2480-05	GPX5	5	1.13	10.53	11.66	10.55	89.5	
Q2480-06	GPX6	6	1.17	10.56	11.73	10.35	86.9	
Q2480-07	GPX7	7	1.18	10.59	11.77	10.48	87.8	
Q2480-08	GPX8	8	1.18	10.37	11.55	10.22	87.2	
Q2484-01	TP-58	9	1.19	10.77	11.96	10.47	86.2	
Q2484-02	TP-57	10	1.18	10.22	11.4	9.97	86.0	
Q2484-03	TP-64	11	1.16	10.50	11.66	10.44	88.4	
Q2484-04	TP-107	12	1.13	10.69	11.82	10.09	83.8	
Q2484-05	TP-1006	13	1.14	10.48	11.62	10.35	87.9	
Q2484-06	TP-104	14	1.15	10.29	11.44	10.13	87.3	
Q2486-01	WASTE	15	1.16	10.31	11.47	9.77	83.5	
Q2486-02	VOC	16	1.18	11.32	12.5	10.65	83.7	
Q2486-03	1	17	1.15	10.80	11.95	10.24	84.2	
Q2486-04	2	18	1.13	10.46	11.59	9.74	82.3	
Q2486-05	3	19	1.13	10.78	11.91	10.12	83.4	
Q2486-06	4	20	1.19	10.00	11.19	9.65	84.6	
Q2486-07	5	21	1.19	10.26	11.45	8.79	74.1	
Q2491-01	EO-1-070225	22	1.19	10.51	11.7	10.9	92.4	
Q2491-02	EO-1-070225-E2	23	1.15	10.14	11.29	10.5	92.2	
Q2492-01	VNJ-254-1	24	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q2492-02	VNJ-254-2	25	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q2493-01	WC-11	26	1.13	10.45	11.58	10.4	88.7	
Q2493-02	WC-11-EPH	27	1.19	10.23	11.42	9.72	83.4	
Q2493-03	WC-11-VOC	28	1.14	9.89	11.03	10.2	91.6	

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 7/3/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:35
In Date: 07/02/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:22
Out Date: 07/03/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB136354

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
Q2494-01	Playhouse/Storag	35	1.00	1.00	2.00	2.00	100.0	Caulk sample
Q2494-02	Playhouse/Storag	36	1.00	1.00	2.00	2.00	100.0	chair form
Q2494-03	Garage Seat Foa	37	1.00	1.00	2.00	2.00	100.0	seat form
Q2494-04	Black foam in Sou	38	1.00	1.00	2.00	2.00	100.0	black form
Q2494-05	Kids Room styof	39	1.00	1.00	2.00	2.00	100.0	yellow form
Q2494-06	Kids Room Lower	40	1.00	1.00	2.00	2.00	100.0	form
Q2494-07	Kids Room styof	41	1.00	1.00	2.00	2.00	100.0	styra form
Q2494-08	Dog Beg Foam	42	1.00	1.00	2.00	2.00	100.0	bag form
Q2494-09	Cotton Layer - Ott	43	1.00	1.00	2.00	2.00	100.0	cotton layer
Q2494-10	Foam Ottoman	44	1.00	1.00	2.00	2.00	100.0	form ottoman
Q2494-11	Foyer North Wall	45	1.00	1.00	2.00	2.00	100.0	foyer north wall plug
Q2494-12	NE Stucco Exterio	46	1.00	1.00	2.00	2.00	100.0	stucco exterior
Q2494-13	Master Mattress	47	1.00	1.00	2.00	2.00	100.0	mattress
Q2495-01	Chair Foam Parke	48	1.00	1.00	2.00	2.00	100.0	chair form
Q2495-02	SE Bedroom (3rd	49	1.00	1.00	2.00	2.00	100.0	mattress
Q2495-03	E Bedroom Ceilin	50	1.00	1.00	2.00	2.00	100.0	ceiling plug
Q2495-04	South East Ceilining	51	1.00	1.00	2.00	2.00	100.0	east selling plug
Q2495-05	Master Bathroom	52	1.00	1.00	2.00	2.00	100.0	selling plug
Q2495-06	3rd Fl Hallway N	53	1.00	1.00	2.00	2.00	100.0	hoalway selling
Q2495-07	SE Bedroom Park	54	1.00	1.00	2.00	2.00	100.0	bedrom plug
Q2495-08	Family Room SW	55	1.00	1.00	2.00	2.00	100.0	family plug
Q2495-09	Foyer SE Corner P	56	1.00	1.00	2.00	2.00	100.0	foyer plug
Q2495-10	Green Carpet Pad	57	1.00	1.00	2.00	2.00	100.0	green carpet pad
Q2495-11	Rug in Dining Roo	58	1.00	1.00	2.00	2.00	100.0	rug
Q2495-12	LR Couch Cushion	59	1.00	1.00	2.00	2.00	100.0	couch cushion
Q2495-13	Dining Room W	60	1.00	1.00	2.00	2.00	100.0	wood and foam
Q2495-14	Kitchen Dining E	61	1.00	1.00	2.00	2.00	100.0	kithean rug
Q2495-15	Garage Southwes	62	1.00	1.00	2.00	2.00	100.0	plug

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 7/3/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:35
In Date: 07/02/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:22
Out Date: 07/03/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB136354

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q2495-16	Zack"s Matterss	63	1.00	1.00	2.00	2.00	100.0	mattess
Q2495-17	Zack"s Couch	64	1.00	1.00	2.00	2.00	100.0	zacks couch
Q2496-02	N exterior center	65	1.00	1.00	2.00	2.00	100.0	stucco wall
Q2496-03	Basement south	66	1.00	1.00	2.00	2.00	100.0	wall cavity stucc
Q2496-04	Entertainment ro	67	1.00	1.00	2.00	2.00	100.0	hvac close
Q2496-05	Black mat outside	68	1.00	1.00	2.00	2.00	100.0	outside mat
Q2496-06	Basement bedroom	69	1.00	1.00	2.00	2.00	100.0	bedrom form
Q2496-07	Basement bedroom-1	70	1.00	1.00	2.00	2.00	100.0	window plug
Q2496-08	Basement bedroom-2	71	1.00	1.00	2.00	2.00	100.0	window plug
Q2496-09	Maeve"s room ma	72	1.00	1.00	2.00	2.00	100.0	mattess
Q2496-10	Master bedroom	73	1.00	1.00	2.00	2.00	100.0	rug
Q2496-11	Master bedroom-1	74	1.00	1.00	2.00	2.00	100.0	chair form
Q2496-12	Master bedroom-2	75	1.00	1.00	2.00	2.00	100.0	mattess form
Q2496-13	Master bedroom-3	76	1.00	1.00	2.00	2.00	100.0	office chair form
Q2496-14	SE bedroom brow	77	1.00	1.00	2.00	2.00	100.0	form
Q2496-15	SE bedroom gray	78	1.00	1.00	2.00	2.00	100.0	gray cushion
Q2496-16	Maeves room gre	79	1.00	1.00	2.00	2.00	100.0	chair form
Q2496-17	Master bedroom-4	80	1.00	1.00	2.00	2.00	100.0	plug
Q2496-18	3rd floor roof ent	81	1.00	1.00	2.00	2.00	100.0	plug
Q2496-19	Living room couc	82	1.00	1.00	2.00	2.00	100.0	couch form
Q2496-20	Living room 2 sea	83	1.00	1.00	2.00	2.00	100.0	form
Q2496-21	Office leather cha	84	1.00	1.00	2.00	2.00	100.0	form
Q2496-22	Pink shoe sole	85	1.00	1.00	2.00	2.00	100.0	shoe sole
Q2496-23	Living room rug S	86	1.00	1.00	2.00	2.00	100.0	rug
Q2496-24	Dining room NE c	87	1.00	1.00	2.00	2.00	100.0	plug
Q2497-01	SW bedroom (San	88	1.00	1.00	2.00	2.00	100.0	plug
Q2497-02	Salman s room SE	89	1.00	1.00	2.00	2.00	100.0	plug compossi
Q2497-03	Sanah s room bed	90	1.00	1.00	2.00	2.00	100.0	sanah form

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 7/3/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:35
In Date: 07/02/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:22
Out Date: 07/03/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB136354

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
Q2497-04	Salman s room m	91	1.00	1.00	2.00	2.00	100.0	mattess form
Q2497-05	Master bed mattr	92	1.00	1.00	2.00	2.00	100.0	mattess
Q2497-06	Kaihaan s room m	93	1.00	1.00	2.00	2.00	100.0	form
Q2497-07	Garage foam yog	94	1.00	1.00	2.00	2.00	100.0	garage foam
Q2497-08	Living room west	95	1.00	1.00	2.00	2.00	100.0	form
Q2497-09	Entrance rug	96	1.00	1.00	2.00	2.00	100.0	rug
Q2497-10	south exterior re	97	1.00	1.00	2.00	2.00	100.0	form
Q2497-11	Garage gym floor	98	1.00	1.00	2.00	2.00	100.0	garage foam
Q2497-12	Rug by the back d	99	1.00	1.00	2.00	2.00	100.0	rug
Q2497-13	Guest bed mattre	100	1.00	1.00	2.00	2.00	100.0	mattess
Q2497-14	Garage backyard	101	1.00	1.00	2.00	2.00	100.0	garage foam
Q2497-15	LR SW conner chai	102	1.00	1.00	2.00	2.00	100.0	form
Q2497-16	Living room NE co	103	1.00	1.00	2.00	2.00	100.0	catters
Q2497-17	Dining room SW c	104	1.00	1.00	2.00	2.00	100.0	chair form
Q2497-18	LR SW rug corner	105	1.00	1.00	2.00	2.00	100.0	rug
Q2497-19	LR west wall plug	106	1.00	1.00	2.00	2.00	100.0	plug
Q2497-20	LR NE corner plug	107	1.00	1.00	2.00	2.00	100.0	plug
Q2497-21	office east wall pl	108	1.00	1.00	2.00	2.00	100.0	plug
Q2497-22	Bedroom north w	109	1.00	1.00	2.00	2.00	100.0	plug
Q2497-23	Kitchen ceiling pl	110	1.00	1.00	2.00	2.00	100.0	ceilling plug
Q2498-01	NE Bd bed foam	111	1.00	1.00	2.00	2.00	100.0	bedrom form
Q2498-02	NE Bd pillow foa	112	1.00	1.00	2.00	2.00	100.0	form
Q2498-03	Master bed NE m	113	1.00	1.00	2.00	2.00	100.0	mattess
Q2500-01	X600-B2	29	1.15	10.84	11.99	11.24	93.1	
Q2500-02	X600-S2	30	1.19	10.36	11.55	10.6	90.8	
Q2500-03	X600-B1	31	1.13	10.42	11.55	11.02	94.9	
Q2500-04	X600-S1	32	1.13	10.32	11.45	10.6	91.8	
Q2500-05	X600-DUP1	33	1.14	10.85	11.99	11.3	93.6	



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 7/3/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:35
In Date: 07/02/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:22
Out Date: 07/03/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB136354

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
Q2500-06	X600-S3	34	1.11	10.71	11.82	11.18	94.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$