

LB136366


7470A

INSTRUMENT ID: Cx2

Linear

 μ Abs.:

51326



A= 0.0000e+000

B= 1.9929e-004

C= -1.5993e-002

Rho= 0.9986976

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/Δ)
0.0	0.000	-0.012	-0.012	20	0.000	20	0				
0.2	0.200	0.211	0.011	1139	0.0 %	1139					
2.5	2.500	2.628	0.128	13266	0.0 %	13266					
5.0	5.000	5.046	0.046	25401	0.0 %	25401					
7.5	7.500	7.114	-0.386	35779	0.0 %	35779					
10.0	10.000	10.213	0.213	51326	0.0 %	51326					

0/Δ)

155

55

2

LB136366 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 03 Jul 2025 09:01:01

Type	Sample ID	Extended ID	Conc.	μ Abs.	Units	Type	Std CondMethod	Date
S	0.0 - 1	50	-	20	PPB	Std	0.00007470A	03 Jul 2025 09:18:50
S	0.2 - 1	50.2	-	1139	PPB	Std	0.20007470A	03 Jul 2025 09:21:08
S	2.5 - 1	52.5	-	13266	PPB	Std	2.50007470A	03 Jul 2025 09:23:25
S	5.0 - 1	55	-	25401	PPB	Std	5.00007470A	03 Jul 2025 09:28:24
S	7.5 - 1	57.5	-	35779	PPB	Std	7.50007470A	03 Jul 2025 09:30:41
S	10.0 - 1	57.0	-	51326	PPB	Std	10.00007470A	03 Jul 2025 09:32:59
U	ICV15 - 1	ICV15	3.7231	18762	PPB	SMPL	-7470A	03 Jul 2025 09:38:41
U	ICB15 - 1	ICB15	-0.0730	-286	PPB	SMPL	-7470A	03 Jul 2025 09:40:56
U	CCV48 - 1	CCV48	5.0202	25271	PPB	SMPL	-7470A	03 Jul 2025 09:43:13
U	CCB48 - 1	CCB48	-0.0435	-138	PPB	SMPL	-7470A	03 Jul 2025 09:45:28
U	CRA - 1	CRA	0.1972	1070	PPB	SMPL	-7470A	03 Jul 2025 09:47:46
U	HighStd - 1	HighStd	10.2279	51402	PPB	SMPL	-7470A	03 Jul 2025 09:50:02
U	ChkStd - 1	ChkStd	6.6931	33665	PPB	SMPL	-7470A	03 Jul 2025 09:52:17
U	PB168687BL - 1	PBW	-0.0834	-338	PPB	SMPL	-7470A	03 Jul 2025 09:54:35
U	PB168687BS - 1	LCSW	4.5760	23042	PPB	SMPL	-7470A	03 Jul 2025 09:56:54
U	Q2447-08 - 1	COMP-1	-0.0676	-259	PPB	SMPL	-7470A	03 Jul 2025 09:59:10
U	Q2447-09 - 1	COMP-2	-0.0132	14	PPB	SMPL	-7470A	03 Jul 2025 10:01:29
U	Q2447-10 - 1	LAW-25-0100	-0.0166	-3	PPB	SMPL	-7470A	03 Jul 2025 10:03:46
U	Q2452-04 - 1	TP-5	-0.0156	2	PPB	SMPL	-7470A	03 Jul 2025 10:06:03
U	Q2452-08 - 1	EP-2	-0.0110	25	PPB	SMPL	-7470A	03 Jul 2025 10:08:18
U	CCV49 - 1	CCV49	4.9417	24877	PPB	SMPL	-7470A	03 Jul 2025 10:10:33
U	CCB49 - 1	CCB49	-0.0287	-64	PPB	SMPL	-7470A	03 Jul 2025 10:12:49
U	Q2459-04 - 1	TP-1	-0.0110	25	PPB	SMPL	-7470A	03 Jul 2025 10:15:09
U	Q2469-04 - 1	WC-1	-0.0224	-32	PPB	SMPL	-7470A	03 Jul 2025 10:17:26
U	Q2469-04DUP - 1	WC-1DUP	-0.0150	5	PPB	SMPL	-7470A	03 Jul 2025 10:19:42
U	Q2469-04MS - 1	WC-1MS	4.2117	21214	PPB	SMPL	-7470A	03 Jul 2025 10:21:58
U	Q2469-04MSD - 1	WC-1MSD	4.5178	22750	PPB	SMPL	-7470A	03 Jul 2025 10:26:57
U	PB168652TB - 1	PB168652TB	-0.0668	-255	PPB	SMPL	-7470A	03 Jul 2025 10:29:14
U	Q2469-04LX5 - 1		-0.0146	7	PPB	SMPL	-7470A	03 Jul 2025 10:31:33
U	Q2469-04A - 1		3.9979	20141	PPB	SMPL	-7470A	03 Jul 2025 10:33:50
U	CCV50 - 1	CCV50	4.9738	25038	PPB	SMPL	-7470A	03 Jul 2025 10:36:08
U	CCB50 - 1	CCB50	-0.0311	-76	PPB	SMPL	-7470A	03 Jul 2025 10:38:26