

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136378

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/07/2025	08:10
2	CAL2	1	Water	NA	NA	20.2	7.00	07/07/2025	08:11
3	CAL3	1	Water	NA	NA	20.3	10.02	07/07/2025	08:14
4	ICV	1	Water	NA	NA	20.2	7.02	07/07/2025	08:15
5	CCV1	1	Water	NA	NA	20.2	2.01	07/07/2025	09:00
6	Q2506-01	1	Water	NA	NA	22.3	6.40	07/07/2025	09:11
7	Q2506-02	1	Water	NA	NA	21.9	5.15	07/07/2025	09:20
8	Q2509-01	1	Water	NA	NA	22.6	7.10	07/07/2025	09:33
9	Q2511-01	1	Water	NA	NA	23.1	5.15	07/07/2025	09:40
10	Q2511-01DUP	1	Water	NA	NA	23.3	5.16	07/07/2025	09:42
11	CCV2	1	Water	NA	NA	20.3	12.02	07/07/2025	09:45

8131378

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH W Q2511 WorkList ID : 190550 Department : Wet-Chemistry Date : 07-07-2025 07:40:46

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2506-01	A #60470	Water	pH	Cool 4 deg C	PSEG03	O21	07/03/2025	9040C
Q2506-02	B #60472	Water	pH	Cool 4 deg C	PSEG03	O21	07/03/2025	9040C
Q2509-01	C AUD-25-01110-0111	Water	pH	Cool 4 deg C	PSEG03	O22	07/03/2025	9040C
Q2511-01	D #60266	Water	pH	Cool 4 deg C	PSEG05	O11	07/03/2025	9040C

Date/Time 07/07/25 08:00

Raw Sample Received by: SA WCI

Raw Sample Relinquished by: JD CSM

Date/Time 07/07/25 13:00

Raw Sample Received by: JD CSM

Raw Sample Relinquished by: SA WCI