

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136379

Slope : 98.4

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/07/2025	08:10
2	CAL2	1	Water	NA	NA	20.2	7.00	07/07/2025	08:11
3	CAL3	1	Water	NA	NA	20.3	10.02	07/07/2025	08:14
4	ICV	1	Water	NA	NA	20.2	7.02	07/07/2025	08:15
5	CCV1	1	Water	NA	NA	20.2	2.02	07/07/2025	08:17
6	Q2509-02	1	Solid	20.02	20	21.9	4.27	07/07/2025	08:30
7	Q2510-01	1	Solid	20.03	20	21.4	7.90	07/07/2025	08:37
8	Q2510-01DUP	1	Solid	20.04	20	21.6	7.92	07/07/2025	08:41
9	CCV2	1	Water	NA	NA	20.2	12.02	07/07/2025	08:45

WORKLIST(Hardcopy Internal Chain)

136379

WorkList Name : PH S Q2510

WorkList ID : 190552

Department : Wet-Chemistry

Date : 07-07-2025 07:57:37

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2509-02	AUD-25-0112	Solid	pH	Cool 4 deg C	PSEG03	O22	07/03/2025	9045D
Q2510-01	#63025-A	Solid	pH	Cool 4 deg C	PSEG05	O22	07/03/2025	9045D

Date/Time 07/07/25 08:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 07/07/25 12:45

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]