

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Iwona

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB136386

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	07/07/2025	09:30
2	Q2477-01		1	12	100.00	>212.0	>212.0	07/07/2025	10:00
3	Q2477-01DUP		1	13	100.00	>212.0	>212.0	07/07/2025	10:30
4	Q2506-01		1	12	100.00	>212.0	>212.0	07/07/2025	11:00
5	Q2506-02		1	12	100.00	>212.0	>212.0	07/07/2025	11:30
6	Q2509-01		1	12	100.00	>212.0	>212.0	07/07/2025	12:00
7	Q2511-01		1	13	100.00	>212.0	>212.0	07/07/2025	12:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 136386

WorkList Name : fp-7-7

WorkList ID : 190551

Department : Wet-Chemistry

Date : 07-07-2025 07:55:32

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2477-01	50728	Water	Flash Point	Cool 4 deg C	PSEG03	A53	07/01/2025	1010B
Q2506-01	#60470	Water	Flash Point	Cool 4 deg C	PSEG03	O21	07/03/2025	1010B
Q2506-02	#60472	Water	Flash Point	Cool 4 deg C	PSEG03	O21	07/03/2025	1010B
Q2509-01	AUD-25-0110-0111	Water	Flash Point	Cool 4 deg C	PSEG03	O22	07/03/2025	1010B
Q2511-01	#60266	Water	Flash Point	Cool 4 deg C	PSEG05	O11	07/03/2025	1010B

Date/Time

07/07/25 09:10

Raw Sample Received by:

12 (w)

Raw Sample Relinquished by:

12 (w)

Date/Time

07/07/25 13:15

Raw Sample Received by:

12 (w)

Raw Sample Relinquished by:

12 (w)