

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Iwona

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB136395

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	07/08/2025	13:30
2	Q2481-01		1	12	100.00	>212.0	>212.0	07/08/2025	14:00
3	Q2481-01DUP		1	13	100.00	>212.0	>212.0	07/08/2025	14:30
4	Q2481-02		1	6	42.00	107.6	107.6	07/08/2025	15:00
5	Q2481-03		1	12	100.00	>212.0	>212.0	07/08/2025	15:30
6	Q2481-04		1	12	100.00	>212.0	>212.0	07/08/2025	16:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 136395

WorkList Name : fp-2481

WorkList ID : 190587

Department : Wet-Chemistry

Date : 07-07-2025 10:07:15

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2481-01	CC0627-AL	Water	Flash Point	Cool 4 deg C	ENV160	A13	06/27/2025	1010B
Q2481-02	CC0627-CLOXPL	Water	Flash Point	Cool 4 deg C	ENV160	A13	06/27/2025	1010B
Q2481-03	CC0625-OXBL	Water	Flash Point	Cool 4 deg C	ENV160	A13	06/27/2025	1010B
Q2481-04	CC0627-AOXL	Water	Flash Point	Cool 4 deg C	ENV160	A13	06/27/2025	1010B

Date/Time 07/07/25 13:20
 Raw Sample Received by: 126-01
 Raw Sample Relinquished by: 28609

Date/Time 07/07/25 16:45
 Raw Sample Received by: 28609
 Raw Sample Relinquished by: 126-01