

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Iwona

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB136398

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	07/08/2025	10:30
2	Q2481-05		1	12	100.00	>212.0	>212.0	07/08/2025	11:00
3	Q2481-05DUP		1	12	100.00	>212.0	>212.0	07/08/2025	11:30
4	Q2481-06		1	6	53.00	127.4	127.1	07/08/2025	12:00
5	Q2481-07		1	6	34.00	93.2	92.9	07/08/2025	12:30
6	Q2481-08		1	13	100.00	>212.0	>212.0	07/08/2025	13:00
7	Q2481-09		1	12	100.00	>212.0	>212.0	07/08/2025	13:30
8	Q2481-10		1	13	100.00	>212.0	>212.0	07/08/2025	14:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FLASH POINT+

WorkList ID : 190589

Department : Wet-Chemistry

Date : 07-07-2025 08:52:22

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2481-05	CC0625-NL	Water	Flash Point	Cool 4 deg C	ENVI60	A13	06/27/2025	1010B
Q2481-06	CC0267-OXPL	Water	Flash Point	Cool 4 deg C	ENVI60	A13	06/27/2025	1010B
Q2481-07	CC0627-OXL	Water	Flash Point	Cool 4 deg C	ENVI60	A13	06/27/2025	1010B
Q2481-08	CC0627-CLOXAL	Water	Flash Point	Cool 4 deg C	ENVI60	A13	06/27/2025	1010B
Q2481-09	CC0627-BL	Water	Flash Point	Cool 4 deg C	ENVI60	A13	06/27/2025	1010B
Q2481-10	CC0627-SFBL	Water	Flash Point	Cool 4 deg C	ENVI60	A13	06/27/2025	1010B

Date/Time 07/08/25 10:15
 Raw Sample Received by: 12 (SC)
 Raw Sample Relinquished by: [Signature]

Date/Time 07/08/25 14:40
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: 12 (SC)