

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB136423

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	07/10/2025	09:00
2	Q2535-01		1	14	100.00	>212.0	>212.0	07/10/2025	09:30
3	Q2535-02		1	12	100.00	>212.0	>212.0	07/10/2025	10:00
4	Q2540-01		1	15	100.00	>212.0	>212.0	07/10/2025	10:30
5	Q2545-01		1	14	100.00	>212.0	>212.0	07/10/2025	11:00
6	Q2550-01		1	14	100.00	>212.0	>212.0	07/10/2025	11:30
7	Q2551-01		1	13	100.00	>212.0	>212.0	07/10/2025	12:00
8	Q2551-01DUP		1	13	100.00	>212.0	>212.0	07/10/2025	12:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06