

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136430

Slope : 98.5

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	07/11/2025	08:40
2	CAL2	1	Water	NA	NA	20.2	7.00	07/11/2025	08:41
3	CAL3	1	Water	NA	NA	20.2	10.02	07/11/2025	08:42
4	ICV	1	Water	NA	NA	20.1	7.01	07/11/2025	08:45
5	CCV1	1	Water	NA	NA	20.2	2.02	07/11/2025	08:50
6	Q2556-01	1	Solid	20.02	20	22.3	11.53	07/11/2025	09:20
7	Q2556-01DUP	1	Solid	20.03	20	22.6	11.55	07/11/2025	09:25
8	Q2561-01	1	Solid	20.02	20	21.6	7.29	07/11/2025	09:37
9	Q2561-02	1	Solid	20.03	20	21.5	5.46	07/11/2025	09:44
10	Q2561-04	1	Solid	20.02	20	22.3	6.49	07/11/2025	10:00
11	CCV2	1	Water	NA	NA	20.2	12.02	07/11/2025	10:10

WORKLIST(Hardcopy Internal Chain)

136430

WorkList Name : ph s q2556

WorkList ID : 190651

Department : Wet-Chemistry

Date : 07-11-2025 07:40:18

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2556-01	RT3997	Solid	pH	Cool 4 deg C	PSEG03	--Sele	07/10/2025	9045D
Q2561-01	AUD-25-0115	Solid	pH	Cool 4 deg C	PSEG03	O21	07/10/2025	9045D
Q2561-02	AUD-25-0116	Solid	pH	Cool 4 deg C	PSEG03	O21	07/10/2025	9045D
Q2561-04	AUD-25-0067	Solid	pH	Cool 4 deg C	PSEG03	O21	07/10/2025	9045D

Date/Time 07-11-25 07:45

Raw Sample Received by: *SAU001*

Raw Sample Relinquished by: *CP*

Date/Time 07-11-25

Raw Sample Received by: *CP*

Raw Sample Relinquished by: *SAU001*