

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136431

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/11/2025	09:00
2	CAL2	1	Water	NA	NA	20.2	7.02	07/11/2025	09:10
3	CAL3	1	Water	NA	NA	20.2	10.02	07/11/2025	09:11
4	ICV	1	Water	NA	NA	20.2	7.00	07/11/2025	09:15
5	CCV1	1	Water	NA	NA	20.2	2.01	07/11/2025	09:20
6	Q2565-01	1	Water	NA	NA	20.4	6.86	07/11/2025	09:30
7	Q2565-01DUP	1	Water	NA	NA	20.5	6.87	07/11/2025	09:32
8	CCV2	1	Water	NA	NA	20.2	12.02	07/11/2025	09:35

WORKLIST(Hardcopy Internal Chain)

136431

WorkList Name : ph w q2565 WorkList ID : 190650 Department : Wet-Chemistry Date : 07-11-2025 07:39:32

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2565-01	MOO-25-0192-0193	Water	pH	Cool 4 deg C	PSEG03		07/10/2025	9040C

Date/Time 07-11-25 07:46
 Raw Sample Received by: SAWCI
 Raw Sample Relinquished by: CP Son

Date/Time 07-11-25 11:00
 Raw Sample Received by: CP Son
 Raw Sample Relinquished by: SAWCI