

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Iwona

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB136457

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	29.00	84.2	84.2	07/14/2025	09:45
2	Q2557-02		1	12	100.00	>212.0	>212.0	07/14/2025	10:15
3	Q2557-02DUP		1	12	100.00	>212.0	>212.0	07/14/2025	10:45
4	Q2565-01		1	13	100.00	>212.0	>212.0	07/14/2025	11:15
5	Q2591-04		1	12	100.00	>212.0	>212.0	07/14/2025	11:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6136457

WorkList Name : FP-07-11

WorkList ID : 190672

Department : Wet-Chemistry

Date : 07-11-2025 11:43:58

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2557-02	OILY WATER	Water	Flash Point	Cool 4 deg C	PSEG03	O21	07/10/2025	1010B
Q2565-01	MOO-25-0192-0193	Water	Flash Point	Cool 4 deg C	PSEG03		07/10/2025	1010B
Q2591-04	NWB-2187	Water	Flash Point	Cool 4 deg C	PSEG03	D41	07/11/2025	1010B

Date/Time 07/11/2025 04:30
 Raw Sample Received by: 12(WC)
 Raw Sample Relinquished by: 00(WC)

Date/Time 07/11/2025 12:50
 Raw Sample Received by: 00(WC)
 Raw Sample Relinquished by: 12(WC)