

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136473

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/15/2025	09:00
2	CAL2	1	Water	NA	NA	20.2	7.00	07/15/2025	09:01
3	CAL3	1	Water	NA	NA	20.2	10.02	07/15/2025	09:05
4	ICV	1	Water	NA	NA	20.3	7.00	07/15/2025	09:11
5	CCV1	1	Water	NA	NA	20.2	2.02	07/15/2025	09:15
6	Q2596-01	1	Water	NA	NA	24.5	7.34	07/15/2025	09:30
7	Q2602-01	1	Water	NA	NA	22.4	10.20	07/15/2025	09:40
8	Q2602-01DUP	1	Water	NA	NA	22.6	10.21	07/15/2025	09:42
9	CCV2	1	Water	NA	NA	20.3	12.02	07/15/2025	09:45

WORKLIST(Hardcopy Internal Chain)

JB 136473

WorkList Name : ph q2602 WorkList ID : 190716 Department : Wet-Chemistry Date : 07-15-2025 08:19:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2596-01	400025	Water	pH	Cool 4 deg C	PSEG03	D41	07/14/2025	9040C
Q2602-01	FRAC-TANK-266380	Water	pH	Cool 4 deg C	PSEG03	D31	07/14/2025	9040C

Date/Time 07-15-25 08:25
 Raw Sample Received by: JB wet
 Raw Sample Relinquished by: JB wet

Date/Time 07-15-25 13:00
 Raw Sample Received by: JB wet
 Raw Sample Relinquished by: JB wet