

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Eman

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB136492

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	07/16/2025	11:30
2	Q2608-01		1	15	100.00	>212.0	>212.0	07/16/2025	12:00
3	Q2608-01DUP		1	15	100.00	>212.0	>212.0	07/16/2025	12:30
4	Q2608-05		1	14	100.00	>212.0	>212.0	07/16/2025	13:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

10136492

WorkList Name : FP-071625

WorkList ID : 190754

Department : Wet-Chemistry

Date : 07-16-2025 08:20:06

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2608-01	A2028	Water	Flash Point	Cool 4 deg C	PSEG03	D41	07/15/2025	1010B
Q2608-05	60265	Water	Flash Point	Cool 4 deg C	PSEG03	D41	07/15/2025	1010B

Date/Time 07/16/25 11:26
Raw Sample Received by: EM(WC)
Raw Sample Relinquished by: JPC(250)

Date/Time 07/16/25 13:40
Raw Sample Received by: JPC(250)
Raw Sample Relinquished by: EM(WC)