

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB136495

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/16/2025	08:00
2	CAL2	1	Water	NA	NA	20.2	7.00	07/16/2025	08:03
3	CAL3	1	Water	NA	NA	20.3	10.02	07/16/2025	08:05
4	ICV	1	Water	NA	NA	20.2	7.00	07/16/2025	08:11
5	CCV1	1	Water	NA	NA	20.2	2.01	07/16/2025	08:15
6	Q2605-01	1	Solid	20.02	20	22.6	8.72	07/16/2025	08:25
7	Q2605-01DUP	1	Solid	20.03	20	22.7	8.73	07/16/2025	08:26
8	Q2605-02	1	Solid	20.04	20	23.2	8.99	07/16/2025	08:35
9	Q2605-03	1	Solid	20.02	20	23.1	8.93	07/16/2025	08:47
10	Q2605-04	1	Solid	20.03	20	23.3	9.33	07/16/2025	08:55
11	Q2608-04	1	Solid	20.02	20	24.2	6.02	07/16/2025	09:10
12	Q2609-02	1	Solid	20.01	20	23.1	7.12	07/16/2025	09:20
13	Q2609-06	1	Solid	20.03	20	22.8	6.12	07/16/2025	09:27
14	Q2614-06	1	Solid	20.02	20	23.2	7.48	07/16/2025	09:35
15	CCV2	1	Water	NA	NA	20.2	12.02	07/16/2025	09:40

WORKLIST(Hardcopy Internal Chain)

0716145

WorkList Name : corrosivity q2614 WorkList ID : 190746 Department : Wet-Chemistry Date : 07-16-2025 07:30:56

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2605-01	V908	Solid	Corrosivity	Cool 4 deg C	PSEG03	D41	07/15/2025	9045D
Q2605-02	VB16135	Solid	Corrosivity	Cool 4 deg C	PSEG03	D41	07/15/2025	9045D
Q2605-03	VB15061	Solid	Corrosivity	Cool 4 deg C	PSEG03	D41	07/15/2025	9045D
Q2605-04	V897	Solid	Corrosivity	Cool 4 deg C	PSEG03	D41	07/15/2025	9045D
Q2608-04	60271	Solid	Corrosivity	Cool 4 deg C	PSEG03	D41	07/15/2025	9045D
Q2609-02	710-ABC	Solid	Corrosivity	Cool 4 deg C	PSEG03	D41	07/15/2025	9045D
Q2609-06	709-ABC	Solid	Corrosivity	Cool 4 deg C	PSEG03	D41	07/15/2025	9045D
Q2614-06	HR-MCN-COMP-01	Solid	Corrosivity	Cool 4 deg C	PSEG03	D41	07/15/2025	9045D

Date/Time 07/16/25 07:45

Raw Sample Received by: SA WPC

Raw Sample Relinquished by: SA WPC

Date/Time 07/16/25 11:30

Raw Sample Received by: SA WPC

Raw Sample Relinquished by: SA WPC