

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136497

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/16/2025	08:30
2	CAL2	1	Water	NA	NA	20.2	7.02	07/16/2025	08:31
3	CAL3	1	Water	NA	NA	20.2	10.02	07/16/2025	08:33
4	ICV	1	Water	NA	NA	20.2	7.01	07/16/2025	08:37
5	CCV1	1	Water	NA	NA	20.2	2.01	07/16/2025	08:44
6	Q2608-01	1	Water	NA	NA	24.2	7.76	07/16/2025	08:50
7	Q2608-01DUP	1	Water	NA	NA	24.4	7.79	07/16/2025	09:00
8	Q2608-05	1	Water	NA	NA	27.9	6.81	07/16/2025	09:15
9	Q2608-05DUP	1	Water	NA	NA	27.6	6.82	07/16/2025	09:19
10	CCV2	1	Water	NA	NA	20.3	12.02	07/16/2025	09:20

WORKLIST(Hardcopy Internal Chain)

18 AB 1376495 1313649# 13136497

WorkList Name : ph w q2608

WorkList ID : 190745

Department : Wet-Chemistry

Date : 07-16-2025 07:29:58

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2608-01	A2028	Water	pH	Cool 4 deg C	PSEG03	D41	07/15/2025	9040C
Q2608-05	60265	Water	pH	Cool 4 deg C	PSEG03	D41	07/15/2025	9040C

Date/Time 07/16/25 07:40

Raw Sample Received by: SA (w/c)

Raw Sample Relinquished by: SA (w/c)

Date/Time 07/16/25

Raw Sample Received by: SA (w/c)

Raw Sample Relinquished by: SA (w/c)