

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136509

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/16/2025	16:10
2	CAL2	1	Water	NA	NA	20.2	7.00	07/16/2025	16:11
3	CAL3	1	Water	NA	NA	20.2	10.02	07/16/2025	16:15
4	ICV	1	Water	NA	NA	20.2	7.00	07/16/2025	16:20
5	CCV1	1	Water	NA	NA	20.2	2.01	07/16/2025	16:22
6	Q2623-01	1	Water	NA	NA	23.6	6.60	07/16/2025	16:30
7	Q2623-01DUP	1	Water	NA	NA	23.7	6.61	07/16/2025	16:32
8	Q2623-02	1	Water	NA	NA	23.9	6.70	07/16/2025	16:38
9	Q2623-03	1	Water	NA	NA	23.7	8.50	07/16/2025	16:40
10	Q2624-01	1	Water	NA	NA	26.6	7.13	07/16/2025	16:50
11	Q2624-02	1	Water	NA	NA	26.4	7.10	07/16/2025	16:55
12	Q2624-03	1	Water	NA	NA	26.7	7.05	07/16/2025	17:00
13	CCV2	1	Water	NA	NA	20.2	12.02	07/16/2025	17:02

WORKLIST(Hardcopy Internal Chain)

136509

WorkList Name : ph w q2623

WorkList ID : 190780

Department : Wet-Chemistry

Date : 07-16-2025 15:41:14

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2623-01	3362	Water	pH	Cool 4 deg C	PSEG03	D41	07/16/2025	9040C
Q2623-02	3363	Water	pH	Cool 4 deg C	PSEG03	D41	07/16/2025	9040C
Q2623-03	3364	Water	pH	Cool 4 deg C	PSEG03	D41	07/16/2025	9040C
Q2624-01	449	Water	pH	Cool 4 deg C	PSEG03	D41	07/16/2025	9040C
Q2624-02	450	Water	pH	Cool 4 deg C	PSEG03	D41	07/16/2025	9040C
Q2624-03	451	Water	pH	Cool 4 deg C	PSEG03	D41	07/16/2025	9040C

Date/Time

07/16/25 16:00

Raw Sample Received by:

sgwc

Raw Sample Relinquished by:

RJ (EXT-106)

Date/Time

07/16/25

Raw Sample Received by:

RJ (EXT-106)

Raw Sample Relinquished by:

sgwc