


L9186522

7471B

INSTRUMENT ID: CV1

μ Abs.:

49447



Linear

A= 0.0000e+000

B= 1.9874e-004

C= -4.9312e-002

Rho= 0.9992346

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.009	-0.009	203	0.000	203					1
0.2	0.200	0.171	-0.029	1109	0.0 %	1109					15
2.5	2.500	2.436	-0.064	12507	0.0 %	12507					3
5.0	5.000	5.066	0.066	25741	0.0 %	25741					1
7.5	7.500	7.757	0.257	39281	0.0 %	39281					3
10.0	10.000	9.778	-0.222	49447	0.0 %	49447					2

LB136522 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 17 Jul 2025 12:43:22

Sample ID	Extended ID	μ Abs	Conc.	Std Cond	Method	Units	Date	Type	Type	B	C	Rho
0.0 - 1	50	203	-	0.000	7471B	PPB	17 Jul 2025 13:33:43	S	Std			
0.2 - 1	50-2	1109	-	0.200	7471B	PPB	17 Jul 2025 13:36:01	S	Std			
2.5 - 1	52-5	12507	-	2.500	7471B	PPB	17 Jul 2025 13:38:18	S	Std			
5.0 - 1	55	25741	-	5.000	7471B	PPB	17 Jul 2025 13:40:35	S	Std			
7.5 - 1	55-5	39281	-	7.500	7471B	PPB	17 Jul 2025 13:42:54	S	Std			
10.0 - 1	55-5	49447	-	10.000	7471B	PPB	17 Jul 2025 13:50:44	S	Std			
ICV32 - 1	ICV32	20279	3.9809	-	7471B	PPB	17 Jul 2025 13:53:40	U	SMPL	1.98739e-004	-4.93119e-002	9.99235e-001
ICB32 - 1	ICB32	-49	-0.0591	-	7471B	PPB	17 Jul 2025 13:55:57	U	SMPL	1.98739e-004	-4.93119e-002	9.99235e-001
CCV22 - 1	CCV22	26313	5.1801	-	7471B	PPB	17 Jul 2025 13:58:16	U	SMPL	1.98739e-004	-4.93119e-002	9.99235e-001
CCB22 - 1	CCB22	-58	-0.0608	-	7471B	PPB	17 Jul 2025 14:00:31	U	SMPL	1.98739e-004	-4.93119e-002	9.99235e-001
CRA - 1	CRA	1115	0.1723	-	7471B	PPB	17 Jul 2025 14:02:51	U	SMPL	1.98739e-004	-4.93119e-002	9.99235e-001
HighStd - 1	HighStd	50142	9.9159	-	7471B	PPB	17 Jul 2025 14:05:06	U	SMPL	1.98739e-004	-4.93119e-002	9.99235e-001
ChkStd - 1	ChkStd	35727	7.0510	-	7471B	PPB	17 Jul 2025 14:07:21	U	SMPL	1.98739e-004	-4.93119e-002	9.99235e-001
PB168894BL - 1	PBS	-196	-0.0883	-	7471B	PPB	17 Jul 2025 14:09:45	U	SMPL	1.98739e-004	-4.93119e-002	9.99235e-001
PB168894BS - 1	LCSS	20442	4.0133	-	7471B	PPB	17 Jul 2025 14:15:03	U	SMPL	1.98739e-004	-4.93119e-002	9.99235e-001
Q2609-01 - 1	710-ABC	19331	3.7925	-	7471B	PPB	17 Jul 2025 14:17:19	U	SMPL	1.98739e-004	-4.93119e-002	9.99235e-001
Q2609-05 - 1	709-AB	5079	0.9601	-	7471B	PPB	17 Jul 2025 14:19:38	U	SMPL	1.98739e-004	-4.93119e-002	9.99235e-001
Q2611-01 - 1	EO-02-071525	4848	0.9142	-	7471B	PPB	17 Jul 2025 14:19:38	U	SMPL	1.98739e-004	-4.93119e-002	9.99235e-001
Q2612-01 - 1	OR-02-071525	2954	0.5378	-	7471B	PPB	17 Jul 2025 14:22:00	U	SMPL	1.98739e-004	-4.93119e-002	9.99235e-001
Q2614-01 - 1	HR-MCN-COMP-01	15639	3.0588	-	7471B	PPB	17 Jul 2025 14:24:15	U	SMPL	1.98739e-004	-4.93119e-002	9.99235e-001
CCV23 - 1	CCV23	25624	5.0432	-	7471B	PPB	17 Jul 2025 14:26:36	U	SMPL	1.98739e-004	-4.93119e-002	9.99235e-001
CCB23 - 1	CCB23	-13	-0.0519	-	7471B	PPB	17 Jul 2025 14:28:51	U	SMPL	1.98739e-004	-4.93119e-002	9.99235e-001
Q2614-01DUP - 1	HR-MCN-COMP-01DUP	16226	3.1754	-	7471B	PPB	17 Jul 2025 14:31:09	U	SMPL	1.98739e-004	-4.93119e-002	9.99235e-001
Q2614-01MS - 1	HR-MCN-COMP-01MS	38386	7.5795	-	7471B	PPB	17 Jul 2025 14:33:28	U	SMPL	1.98739e-004	-4.93119e-002	9.99235e-001
Q2614-01MSD - 1	HR-MCN-COMP-01MSD	38165	7.5356	-	7471B	PPB	17 Jul 2025 14:38:28	U	SMPL	1.98739e-004	-4.93119e-002	9.99235e-001
Q2622-01 - 1	2819	801	0.1099	-	7471B	PPB	17 Jul 2025 14:40:44	U	SMPL	1.98739e-004	-4.93119e-002	9.99235e-001
Q2614-01LX5 - 1		3346	0.6157	-	7471B	PPB	17 Jul 2025 14:43:09	U	SMPL	1.98739e-004	-4.93119e-002	9.99235e-001
Q2614-01A - 1		39282	7.7576	-	7471B	PPB	17 Jul 2025 14:45:27	U	SMPL	1.98739e-004	-4.93119e-002	9.99235e-001
CCV24 - 1	CCV24	26143	5.1463	-	7471B	PPB	17 Jul 2025 14:47:43	U	SMPL	1.98739e-004	-4.93119e-002	9.99235e-001
CCB24 - 1	CCB24	126	-0.0243	-	7471B	PPB	17 Jul 2025 14:50:00	U	SMPL	1.98739e-004	-4.93119e-002	9.99235e-001
							17 Jul 2025 14:52:20	U	SMPL	1.98739e-004	-4.93119e-002	9.99235e-001