

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB136531

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	07/17/2025	09:20
2	Q2620-02		1	16	100.00	>212.0	>212.0	07/17/2025	09:50
3	Q2620-02DUP		1	16	100.00	>212.0	>212.0	07/17/2025	10:20
4	Q2623-01		1	14	100.00	>212.0	>212.0	07/17/2025	10:50
5	Q2623-02		1	14	100.00	>212.0	>212.0	07/17/2025	11:20
6	Q2623-03		1	15	100.00	>212.0	>212.0	07/17/2025	11:50
7	Q2624-01		1	13	100.00	>212.0	>212.0	07/17/2025	12:20
8	Q2624-02		1	15	100.00	>212.0	>212.0	07/17/2025	12:50
9	Q2624-03		1	12	100.00	>212.0	>212.0	07/17/2025	13:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66136531

WorkList Name : fp 7-17

WorkList ID : 190789

Department : Wet-Chemistry

Date : 07-17-2025 08:20:10

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2620-02	GAS-BUR-0058	Water	Flash Point	Cool 4 deg C	PSEG03	D31	07/16/2025	1010B
Q2623-01	3362	Water	Flash Point	Cool 4 deg C	PSEG03	D41	07/16/2025	1010B
Q2623-02	3363	Water	Flash Point	Cool 4 deg C	PSEG03	D41	07/16/2025	1010B
Q2623-03	3364	Water	Flash Point	Cool 4 deg C	PSEG03	D41	07/16/2025	1010B
Q2624-01	449	Water	Flash Point	Cool 4 deg C	PSEG03	D41	07/16/2025	1010B
Q2624-02	450	Water	Flash Point	Cool 4 deg C	PSEG03	D41	07/16/2025	1010B
Q2624-03	451	Water	Flash Point	Cool 4 deg C	PSEG03	D41	07/16/2025	1010B

Date/Time 07/17/2025 09:10
 Raw Sample Received by: RM (W.C.)
 Raw Sample Relinquished by: P(OC)

Date/Time 07/17/2025 14:20
 Raw Sample Received by: P(OC)
 Raw Sample Relinquished by: R(OC)