

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB136535

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

True Value of CCV3 = 2.00 Control Limits[+/- 0.05].

True Value of CCV4 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/18/2025	08:00
2	CAL2	1	Water	NA	NA	20.2	7.00	07/18/2025	08:02
3	CAL3	1	Water	NA	NA	20.2	10.02	07/18/2025	08:03
4	ICV	1	Water	NA	NA	20.2	7.01	07/18/2025	08:05
5	CCV1	1	Water	NA	NA	20.2	2.01	07/18/2025	08:10
6	Q2633-02	1	Water	NA	NA	20.7	7.25	07/18/2025	08:15
7	Q2633-02DUP	1	Water	NA	NA	20.9	7.26	07/18/2025	08:17
8	Q2636-04	1	Water	NA	NA	21.9	5.31	07/18/2025	08:22
9	Q2636-04DUP	1	Water	NA	NA	22.1	5.33	07/18/2025	08:24
10	Q2636-05	1	Water	NA	NA	20.8	5.26	07/18/2025	08:26
11	Q2636-05DUP	1	Water	NA	NA	20.6	5.28	07/18/2025	08:30
12	Q2636-06	1	Water	NA	NA	22.7	5.41	07/18/2025	08:35
13	Q2636-06DUP	1	Water	NA	NA	23.5	5.43	07/18/2025	08:37
14	Q2636-07	1	Water	NA	NA	22.0	5.17	07/18/2025	08:40
15	Q2636-07DUP	1	Water	NA	NA	22.2	5.19	07/18/2025	08:42
16	CCV2	1	Water	NA	NA	20.2	12.02	07/18/2025	08:45
17	Q2636-08	1	Water	NA	NA	20.7	5.30	07/18/2025	08:50
18	Q2636-08DUP	1	Water	NA	NA	20.9	5.32	07/18/2025	08:55
19	Q2636-09	1	Water	NA	NA	20.8	5.19	07/18/2025	09:00
20	Q2636-09DUP	1	Water	NA	NA	20.7	5.21	07/18/2025	09:03

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	Q2636-10	1	Water	NA	NA	22.1	5.06	07/18/2025	09:05
22	Q2636-10DUP	1	Water	NA	NA	22.2	5.08	07/18/2025	09:09
23	Q2636-11	1	Water	NA	NA	22.7	5.47	07/18/2025	09:11
24	Q2636-11DUP	1	Water	NA	NA	22.9	5.49	07/18/2025	09:14
25	Q2636-12	1	Water	NA	NA	22.7	5.65	07/18/2025	09:22
26	Q2636-12DUP	1	Water	NA	NA	22.8	5.67	07/18/2025	09:24
27	CCV3	1	Water	NA	NA	20.2	2.01	07/18/2025	09:30
28	Q2636-13	1	Water	NA	NA	21.6	5.17	07/18/2025	09:35
29	Q2636-13DUP	1	Water	NA	NA	21.7	5.19	07/18/2025	09:40
30	Q2636-14	1	Water	NA	NA	21.2	5.36	07/18/2025	09:47
31	Q2636-14DUP	1	Water	NA	NA	21.3	5.38	07/18/2025	09:50
32	Q2636-15	1	Water	NA	NA	21.5	5.80	07/18/2025	09:55
33	Q2636-15DUP	1	Water	NA	NA	21.6	5.82	07/18/2025	10:00
34	CCV4	1	Water	NA	NA	20.2	12.02	07/18/2025	10:05

WORKLIST(Hardcopy Internal Chain)

136535

WorkList Name : ph q2633

WorkList ID : 190812

Department : Wet-Chemistry

Date : 07-18-2025 07:33:04

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2633-02	✓ FIBER-GLASS-TANK	Water	pH	Cool 4 deg C	PSEG03	D41	07/17/2025	9040C
Q2636-04	0715-A	Water	pH	Cool 4 deg C	PSEG03	D41	07/17/2025	9040C
Q2636-05	0715-B	Water	pH	Cool 4 deg C	PSEG03	D41	07/17/2025	9040C
Q2636-06	0715-C	Water	pH	Cool 4 deg C	PSEG03	D41	07/17/2025	9040C
Q2636-07	0715-D	Water	pH	Cool 4 deg C	PSEG03	D41	07/17/2025	9040C
Q2636-08	0715-E	Water	pH	Cool 4 deg C	PSEG03	D41	07/17/2025	9040C
Q2636-09	0715-F	Water	pH	Cool 4 deg C	PSEG03	D41	07/17/2025	9040C
Q2636-10	0715-G	Water	pH	Cool 4 deg C	PSEG03	D41	07/17/2025	9040C
Q2636-11	0715-H	Water	pH	Cool 4 deg C	PSEG03	D41	07/17/2025	9040C
Q2636-12	0715-I	Water	pH	Cool 4 deg C	PSEG03	D41	07/17/2025	9040C
Q2636-13	0715-J	Water	pH	Cool 4 deg C	PSEG03	D41	07/17/2025	9040C
Q2636-14	0715-K	Water	pH	Cool 4 deg C	PSEG03	D41	07/17/2025	9040C
Q2636-15	0715-L	Water	pH	Cool 4 deg C	PSEG03	D41	07/17/2025	9040C

Date/Time 07/18/25 07:45

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 07/18/25

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

13:30