

LB136538


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

45612



A= 0.0000e+000

B= 2.1628e-004

C= -1.8563e-002

Rho= 0.9995380

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.021	0.021	182	0.000	182	0				-
0.2	0.200	0.161	-0.039	831	0.0 %	831					-20
2.5	2.500	2.394	-0.106	11153	0.0 %	11153					-4
5.0	5.000	5.140	0.140	23853	0.0 %	23853					3
7.5	7.500	7.638	0.138	35399	0.0 %	35399					2
10.0	10.000	9.846	-0.154	45612	0.0 %	45612					-2

LB136538 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 Jul 2025 10:58:57

Type	Sample ID	Extended ID	Conc.	μ Abs.	Units	Type	Std Conc	Method	Date
S	0.0 - 1	50	-	182	PPB	Std	0.0000	7470A	18 Jul 2025 11:12:16
S	0.2 - 1	50.2	-	831	PPB	Std	0.2000	7470A	18 Jul 2025 11:14:33
S	2.5 - 1	52.5	-	11153	PPB	Std	2.5000	7470A	18 Jul 2025 11:16:51
S	5.0 - 1	55	-	23853	PPB	Std	5.0000	7470A	18 Jul 2025 11:21:50
S	7.5 - 1	57.5	-	35399	PPB	Std	7.5000	7470A	18 Jul 2025 11:24:07
S	10.0 - 1	510	-	45612	PPB	Std	10.0000	7470A	18 Jul 2025 11:29:16
U	ICV33 - 1	ICV33	4.0064	18610	PPB	SMPL	-7470A		18 Jul 2025 11:32:12
U	ICB33 - 1	ICB33	-0.0532	-160	PPB	SMPL	-7470A		18 Jul 2025 11:34:27
U	CCV25 - 1	CCV25	5.1356	23831	PPB	SMPL	-7470A		18 Jul 2025 11:36:45
U	CCB25 - 1	CCB25	-0.0876	-319	PPB	SMPL	-7470A		18 Jul 2025 11:39:00
U	CRA - 1	CRA	0.2261	1131	PPB	SMPL	-7470A		18 Jul 2025 11:41:17
U	HighStd - 1	HighStd	9.7210	45032	PPB	SMPL	-7470A		18 Jul 2025 11:43:32
U	ChkStd - 1	ChkStd	6.7941	31499	PPB	SMPL	-7470A		18 Jul 2025 11:45:47
U	PB168915BL - 1	PBW	-0.0912	-336	PPB	SMPL	-7470A		18 Jul 2025 11:48:06
U	PB168915BS - 1	LCSW	3.9623	18406	PPB	SMPL	-7470A		18 Jul 2025 11:53:17
U	Q2620-02DLX10 - 1	GAS-BUR-25-0058	2.9741	13837	PPB	SMPL	-7470A		18 Jul 2025 11:55:33
U	Q2623-01 - 1	3362	0.0013	92	PPB	SMPL	-7470A		18 Jul 2025 11:57:52
U	Q2623-02 - 1	3363	0.1677	861	PPB	SMPL	-7470A		18 Jul 2025 12:00:11
U	Q2623-03 - 1	3364	2.8744	13376	PPB	SMPL	-7470A		18 Jul 2025 12:02:29
U	Q2624-01 - 1	449	-0.1172	-456	PPB	SMPL	-7470A		18 Jul 2025 12:04:46
U	CCV26 - 1	CCV26	4.8084	22318	PPB	SMPL	-7470A		18 Jul 2025 12:07:19
U	CCB26 - 1	CCB26	-0.0484	-138	PPB	SMPL	-7470A		18 Jul 2025 12:09:35
U	Q2624-02 - 1	450	0.0195	176	PPB	SMPL	-7470A		18 Jul 2025 12:11:52
U	Q2624-03 - 1	451	0.0247	200	PPB	SMPL	-7470A		18 Jul 2025 12:14:07
U	Q2633-02 - 1	FIBER-GLASS-TANK	-0.0155	14	PPB	SMPL	-7470A		18 Jul 2025 12:16:22
U	Q2633-02DUP - 1	FIBER-GLASS-TANKDUP	-0.0145	19	PPB	SMPL	-7470A		18 Jul 2025 12:18:38
U	Q2633-02MS - 1	FIBER-GLASS-TANKMS	2.5020	11654	PPB	SMPL	-7470A		18 Jul 2025 12:20:54
U	Q2633-02MSD - 1	FIBER-GLASS-TANKMSD	2.2844	10648	PPB	SMPL	-7470A		18 Jul 2025 12:23:11
U	Q2633-02LX5 - 1		-0.0283	-45	PPB	SMPL	-7470A		18 Jul 2025 12:25:28
U	Q2633-02A - 1		3.7456	17404	PPB	SMPL	-7470A		18 Jul 2025 12:27:46
U	CCV27 - 1	CCV27	4.7362	21984	PPB	SMPL	-7470A		18 Jul 2025 12:30:03
U	CCB27 - 1	CCB27	-0.0670	-224	PPB	SMPL	-7470A		18 Jul 2025 12:32:21