

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Eman

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB136540

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.7	07/18/2025	09:20
2	Q2633-02		1	15	100.00	>212.0	>212.0	07/18/2025	09:50
3	Q2633-02DUP		1	15	100.00	>212.0	>212.0	07/18/2025	11:00
4	Q2636-01		1	13	100.00	>212.0	>212.0	07/18/2025	11:30
5	Q2636-02		1	14	100.00	>212.0	>212.0	07/18/2025	12:00
6	Q2636-03		1	12	100.00	>212.0	>212.0	07/18/2025	12:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16136540

WorkList Name : FP-071825

WorkList ID : 190815

Department : Wet-Chemistry

Date : 07-18-2025 08:07:54

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2633-02	FIBER-GLASS-TANK	Water	Flash Point	Cool 4 deg C	PSEG03	D41	07/17/2025	1010B
Q2636-01	COMP-0715-A-0715-E	Water	Flash Point	Cool 4 deg C	PSEG03	D41	07/17/2025	1010B
Q2636-02	COMP-0715-F-0715-J	Water	Flash Point	Cool 4 deg C	PSEG03	D41	07/17/2025	1010B
Q2636-03	COMP-0715-K-0716-L	Water	Flash Point	Cool 4 deg C	PSEG03	D41	07/17/2025	1010B

Date/Time 07/18/25 09:10
Raw Sample Received by: EM(USE)
Raw Sample Relinquished by: [Signature]

Date/Time 07/18/25 13:10
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: EM(WC)