

Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: COD

SUPERVISOR REVIEW BY: jignesh

Run Number: LB136567

Reagent/Standard	Lot/Log #
COD calibration std. 150 ppm	WP113238
COD calibration std. 100 ppm	WP113237
COD calibration std. 50 ppm	WP113235
COD calibration std. 10 ppm	WP113234
COD calibration std. 0 ppm	WP113233
COD ICV-LCS std, 50ppm	WP113240
COD calibration std. 75 ppm	WP113236
COD CCV std, 50ppm	WP114021
COD ICV-LCS std, 50ppm	WP114022
RL CHECK	WP114023
COD Digestion Vials Low Level 0-150Mg/L	W3228

Temp In(C): 147	Date In: 07/22/2025	Time In: 09:45
Temp Out(C): 151	Date Out: 07/22/2025	Time Out: 11:45

Intercept: 0.8179

Slope: 0.9847

Regression: 0.9995

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	-0.831		05/28/2025	13:10
2	CAL2	10	1	Water	9.000	8.309	-16.9	05/28/2025	13:10
3	CAL3	50	1	Water	52.000	51.977	4	05/28/2025	13:11
4	CAL4	75	1	Water	77.000	77.366	3.2	05/28/2025	13:11
5	CAL5	100	1	Water	99.000	99.708	-0.3	05/28/2025	13:12
6	CAL6	150	1	Water	147.000	148.453	-1	05/28/2025	13:12

Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: COD

SUPERVISOR REVIEW BY: jignesh

Run Number: LB136567

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50	NA	NA	1	Water	51.000	50.962	05/28/2025	13:13
2	ICB		NA	NA	1	Water	0.000	-0.831	05/28/2025	13:13
3	CCV1	50	NA	NA	1	Water	49.000	48.931	07/22/2025	13:00
4	CCB1		NA	NA	1	Water	1.000	0.185	07/22/2025	13:00
5	RL Check	10	NA	NA	1	Water	10.000	9.325	07/22/2025	13:01
6	LB136567BL		NA	NA	1	Water	0.000	-0.831	07/22/2025	13:01
7	LB136567BS	50	NA	NA	1	Water	51.000	50.962	07/22/2025	13:02
8	Q2633-02		NA	NA	1	Water	2.000	1.200	07/22/2025	13:02
9	Q2633-02DUP		NA	NA	1	Water	2.000	1.200	07/22/2025	13:03
10	Q2633-02MS	50	NA	NA	1	Water	49.000	48.931	07/22/2025	13:03
11	Q2633-02MSD	50	NA	NA	1	Water	48.000	47.915	07/22/2025	13:04
12	CCV2	50	NA	NA	1	Water	50.000	49.946	07/22/2025	13:04
13	CCB2		NA	NA	1	Water	1.000	0.185	07/22/2025	13:05

WORKLIST(Hardcopy Internal Chain)

LB136567

WorkList Name : COD-072225 WorkList ID : 190869 Department : Wet-Chemistry Date : 07-22-2025 09:27:56

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2633-02	FIBER-GLASS-TANK	Water	COD	Conc H2SO4 to pH < 2	PSEG03	D41	07/17/2025	SM5220 D

Date/Time 07/22/25 09:30
 Raw Sample Received by: 12 (WC)
 Raw Sample Relinquished by: 28 (WC)

Date/Time 07/22/25 10:20
 Raw Sample Received by: 28 (WC)
 Raw Sample Relinquished by: 12 (WC)