

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB136574

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/22/2025	15:15
2	CAL2	1	Water	NA	NA	20.2	7.00	07/22/2025	15:16
3	CAL3	1	Water	NA	NA	20.3	10.02	07/22/2025	15:19
4	ICV	1	Water	NA	NA	20.2	7.01	07/22/2025	15:22
5	CCV1	1	Water	NA	NA	20.2	2.01	07/22/2025	15:24
6	Q2655-03	1	Water	NA	NA	20.1	5.11	07/22/2025	15:35
7	Q2655-03DUP	1	Water	NA	NA	20.2	5.12	07/22/2025	15:36
8	Q2655-04	1	Water	NA	NA	20.2	5.67	07/22/2025	15:40
9	Q2655-04DUP	1	Water	NA	NA	20.4	5.69	07/22/2025	15:44
10	Q2657-01	1	Water	NA	NA	20.1	8.20	07/22/2025	15:50
11	Q2657-01DUP	1	Water	NA	NA	20.3	8.22	07/22/2025	16:00
12	Q2660-05	1	Water	NA	NA	24.3	7.60	07/22/2025	16:10
13	Q2660-05DUP	1	Water	NA	NA	24.4	7.62	07/22/2025	16:14
14	Q2672-05	1	Water	NA	NA	23.6	5.44	07/22/2025	16:25
15	Q2672-05DUP	1	Water	NA	NA	23.7	5.46	07/22/2025	16:30
16	CCV2	1	Water	NA	NA	20.3	12.02	07/22/2025	16:35

WORKLIST(Hardcopy Internal Chain)

136514

WorkList Name : ph q2655 WorkList ID : 190876 Department : Wet-Chemistry Date : 07-22-2025 12:31:29

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2655-03	71125	Water	pH	Cool 4 deg C	PSEG03	D41	07/21/2025	9040C
Q2655-04	71425	Water	pH	Cool 4 deg C	PSEG03	D41	07/21/2025	9040C
Q2657-01	12825-A	Water	pH	Cool 4 deg C	PSEG03	D41	07/21/2025	9040C
Q2660-05	MOO-25-0220	Water	pH	Cool 4 deg C	PSEG03	D31	07/21/2025	9040C
Q2672-05	AUD-25-0122	Water	pH	Cool 4 deg C	PSEG03	D31	07/22/2025	9040C

Date/Time 07/22/25 15:10
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 07/22/25 18:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]