

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB136578

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	07/23/2025	09:45
2	Q2655-03		1	15	100.00	>212.0	>212.0	07/23/2025	10:15
3	Q2655-03DUP		1	15	100.00	>212.0	>212.0	07/23/2025	10:45
4	Q2655-04		1	13	100.00	>212.0	>212.0	07/23/2025	11:15
5	Q2657-01		1	12	100.00	>212.0	>212.0	07/23/2025	11:45
6	Q2660-05		1	15	100.00	>212.0	>212.0	07/23/2025	12:15
7	Q2672-05		1	14	100.00	>212.0	>212.0	07/23/2025	12:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

db136578

WorkList Name : fp7-22 WorkList ID : 190859 Department : Wet-Chemistry Date : 07-22-2025 08:10:00

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2655-03	71125	Water	Flash Point	Cool 4 deg C	PSEG03	D41	07/21/2025	1010B
Q2655-04	71425	Water	Flash Point	Cool 4 deg C	PSEG03	D41	07/21/2025	1010B
Q2657-01	12825-A	Water	Flash Point	Cool 4 deg C	PSEG03	D41	07/21/2025	1010B
Q2660-05	MOO-25-0220	Water	Flash Point	Cool 4 deg C	PSEG03	D31	07/21/2025	1010B
Q2672-05	AUD-25-0122	Water	Flash Point	Cool 4 deg C	PSEG03	D31	07/22/2025	1010B

Date/Time 07/23/2025 09:30
 Raw Sample Received by: RM Lwy
 Raw Sample Relinquished by: JPL (WD)

Date/Time 07/23/2025 13:30
 Raw Sample Received by: JPL (WD)
 Raw Sample Relinquished by: RM Lwy