

LB136591


7470A

INSTRUMENT ID: CY2

Linear

 μ Abs.:

33320



A= 0.0000e+000

B= 2.9760e-004

C= 7.1401e-002

Rho= 0.9999537

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% ID
0.0	0.000	0.046	0.046	-86	0.000	-86	0				-4
0.2	0.200	0.193	-0.007	408	0.0 %	408					-2
2.5	2.500	2.444	-0.056	7972	0.0 %	7972					0
5.0	5.000	4.986	-0.014	16515	0.0 %	16515					1
7.5	7.500	7.544	0.044	25109	0.0 %	25109					0
10.0	10.000	9.987	-0.013	33320	0.0 %	33320					

LB136591 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 24 Jul 2025 11:51:44

Type	Sample ID	Extended ID	Conc.	μ Abs.	Units	Type	Std ConcMethod	Date
S	0.0 - 1	50	-	-86	PPB	Std	0.00007470A	24 Jul 2025 12:03:19
S	0.2 - 1	50.2	-	408	PPB	Std	0.20007470A	24 Jul 2025 12:05:38
S	2.5 - 1	52.5	-	7972	PPB	Std	2.50007470A	24 Jul 2025 12:07:57
S	5.0 - 1	55	-	16515	PPB	Std	5.00007470A	24 Jul 2025 12:10:15
S	7.5 - 1	57.5	-	25109	PPB	Std	7.50007470A	24 Jul 2025 12:12:32
S	10.0 - 1	510	-	33320	PPB	Std	10.00007470A	24 Jul 2025 12:14:50
U	ICV38 - 1	ICV38	4.0247	13284	PPB	SMPL	-7470A	24 Jul 2025 12:18:18
U	ICB38 - 1	ICB38	-0.0467	-397	PPB	SMPL	-7470A	24 Jul 2025 12:20:33
U	CCV44 - 1	CCV44	4.8797	16157	PPB	SMPL	-7470A	24 Jul 2025 12:22:50
U	CCB44 - 1	CCB44	-0.0670	-465	PPB	SMPL	-7470A	24 Jul 2025 12:25:05
U	CRA - 1	CRA	0.2169	489	PPB	SMPL	-7470A	24 Jul 2025 12:27:23
U	HighStd - 1	HighStd	9.8853	32977	PPB	SMPL	-7470A	24 Jul 2025 12:29:38
U	ChkStd - 1	ChkStd	6.6585	22134	PPB	SMPL	-7470A	24 Jul 2025 12:31:54
U	PB168982BL - 1	PBW	0.0419	-99	PPB	SMPL	-7470A	24 Jul 2025 12:36:55
U	PB168982BS - 1	LCSW	4.2384	14002	PPB	SMPL	-7470A	24 Jul 2025 12:39:11
U	Q2655-02 - 1	SOIL	0.0306	-137	PPB	SMPL	-7470A	24 Jul 2025 12:41:27
U	Q2660-02 - 1	MOO-25-0205	0.0196	-174	PPB	SMPL	-7470A	24 Jul 2025 12:43:46
U	Q2660-04 - 1	MOO-25-0218	0.0491	-75	PPB	SMPL	-7470A	24 Jul 2025 12:46:03
U	Q2668-04 - 1	TP-2	0.0681	-11	PPB	SMPL	-7470A	24 Jul 2025 12:48:20
U	Q2668-08 - 1	TP-3	0.0208	-170	PPB	SMPL	-7470A	24 Jul 2025 12:50:35
U	CCV45 - 1	CCV45	5.0857	16849	PPB	SMPL	-7470A	24 Jul 2025 12:55:25
U	CCB45 - 1	CCB45	-0.0822	-516	PPB	SMPL	-7470A	24 Jul 2025 12:57:41
U	Q2668-12 - 1	TP-1	0.0428	-96	PPB	SMPL	-7470A	24 Jul 2025 13:00:00
U	Q2672-02 - 1	AUD-25-0123-0127	0.0625	-30	PPB	SMPL	-7470A	24 Jul 2025 13:02:15
U	Q2672-04 - 1	AUD-25-0128-0132	0.0973	87	PPB	SMPL	-7470A	24 Jul 2025 13:04:31
U	Q2672-04DUP - 1	AUD-25-0128-0132DUP	0.0660	-18	PPB	SMPL	-7470A	24 Jul 2025 13:06:47
U	Q2672-04MS - 1	AUD-25-0128-0132MS	3.9456	13018	PPB	SMPL	-7470A	24 Jul 2025 13:09:03
U	Q2672-04MSD - 1	AUD-25-0128-0132MSD	3.9518	13039	PPB	SMPL	-7470A	24 Jul 2025 13:11:19
U	PB168953TB - 1	PB168953TB	-0.0643	-456	PPB	SMPL	-7470A	24 Jul 2025 13:13:38
U	Q2672-04LX5 - 1		0.0268	-150	PPB	SMPL	-7470A	24 Jul 2025 13:17:37
U	Q2672-04A - 1		4.5339	14995	PPB	SMPL	-7470A	24 Jul 2025 13:19:54
U	CCV46 - 1	CCV46	5.2452	17385	PPB	SMPL	-7470A	24 Jul 2025 13:27:32
U	CCB46 - 1	CCB46	-0.0467	-397	PPB	SMPL	-7470A	24 Jul 2025 13:29:47