

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136592

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/24/2025	11:47
2	CAL2	1	Water	NA	NA	20.2	7.00	07/24/2025	11:48
3	CAL3	1	Water	NA	NA	20.3	10.02	07/24/2025	11:50
4	ICV	1	Water	NA	NA	20.2	7.01	07/24/2025	12:00
5	CCV1	1	Water	NA	NA	20.3	2.01	07/24/2025	12:05
6	Q2688-12	1	Water	NA	NA	20.1	7.55	07/24/2025	12:15
7	Q2688-12DUP	1	Water	NA	NA	20.2	7.57	07/24/2025	12:19
8	Q2688-13	1	Water	NA	NA	20.6	7.90	07/24/2025	12:28
9	CCV2	1	Water	NA	NA	20.3	12.02	07/24/2025	12:30

WORKLIST(Hardcopy Internal Chain)

136592

WorkList Name : ph w q2688

WorkList ID : 190919

Department : Wet-Chemistry

Date : 07-24-2025 11:32:29

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2688-12	☒ SOM-78	Water	pH	Cool 4 deg C	PSEG03	D21	07/23/2025	9040C
Q2688-13	☒ TOTE-COMP	Water	pH	Cool 4 deg C	PSEG03	D21	07/23/2025	9040C

Date/Time 07/24/25 11:40
 Raw Sample Received by: SA GARCIA
 Raw Sample Relinquished by: SA GARCIA

Date/Time 07/24/25 17:30
 Raw Sample Received by: SA GARCIA
 Raw Sample Relinquished by: SA GARCIA