

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Eman

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB136608

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	10	28.00	82.4	82.1	07/24/2025	12:25
2	Q2688-09		1	14	100.00	>212.0	>212.0	07/24/2025	12:55
3	Q2688-09DUP		1	14	100.00	>212.0	>212.0	07/24/2025	13:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB136608

WorkList Name : FLAS POINT-2688

WorkList ID : 190939

Department : Wet-Chemistry

Date : 07-24-2025 08:04:48

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2688-09	SOM-81	Solid	Flash Point	Cool 4 deg C	PSEG03	D21	07/23/2025	1010B

Date/Time 07/24/2025 12:10
Raw Sample Received by: EM(wc)
Raw Sample Relinquished by: MGR

Date/Time 07/24/2025 14:10
Raw Sample Received by: SP Cee
Raw Sample Relinquished by: EM(wc)