

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136610

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/24/2025	15:27
2	CAL2	1	Water	NA	NA	20.2	7.00	07/24/2025	15:28
3	CAL3	1	Water	NA	NA	20.2	10.02	07/24/2025	15:30
4	ICV	1	Water	NA	NA	20.2	7.01	07/24/2025	15:33
5	CCV1	1	Water	NA	NA	20.1	2.02	07/24/2025	16:15
6	Q2691-03	1	Water	NA	NA	23.7	6.64	07/24/2025	16:25
7	Q2691-03DUP	1	Water	NA	NA	23.9	6.66	07/24/2025	16:30
8	Q2691-04	1	Water	NA	NA	23.9	6.76	07/24/2025	16:37
9	Q2691-04DUP	1	Water	NA	NA	24.1	6.78	07/24/2025	16:42
10	Q2691-05	1	Water	NA	NA	23.7	6.84	07/24/2025	16:50
11	Q2691-05DUP	1	Water	NA	NA	23.8	6.86	07/24/2025	16:55
12	CCV2	1	Water	NA	NA	20.2	12.02	07/24/2025	17:00

WORKLIST(Hardcopy Internal Chain)

U-136610

WorkList Name : PH Q2691

WorkList ID : 190944

Department : Wet-Chemistry

Date : 07-24-2025 15:14:54

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2691-03	B 296	Water	pH	Cool 4 deg C	PSEG03	D21	07/24/2025	9040C
Q2691-04	B 297	Water	pH	Cool 4 deg C	PSEG03	D21	07/24/2025	9040C
Q2691-05	B 298	Water	pH	Cool 4 deg C	PSEG03	D21	07/24/2025	9040C

Date/Time 07/24/25 15120

Raw Sample Received by: SB WOC

Raw Sample Relinquished by: CP R

Date/Time 07/24/25

Raw Sample Received by: CP R

Raw Sample Relinquished by: SB WOC

17150

CP R

SB WOC