

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Eman

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB136619

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	07/25/2025	11:15
2	Q2691-03		1	15	100.00	>212.0	>212.0	07/25/2025	11:45
3	Q2691-03DUP		1	15	100.00	>212.0	>212.0	07/25/2025	12:15
4	Q2691-04		1	13	100.00	>212.0	>212.0	07/25/2025	12:45
5	Q2691-05		1	14	100.00	>212.0	>212.0	07/25/2025	13:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6136619

WorkList Name : fp-7-25

WorkList ID : 190956

Department : Wet-Chemistry

Date : 07-25-2025 10:01:19

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2691-03	296	Water	Flash Point	Cool 4 deg C	PSEG03	D21	07/24/2025	1010B
Q2691-04	297	Water	Flash Point	Cool 4 deg C	PSEG03	D21	07/24/2025	1010B
Q2691-05	298	Water	Flash Point	Cool 4 deg C	PSEG03	D21	07/24/2025	1010B

Date/Time 07/25/2025 11.00
 Raw Sample Received by: EM CWC
 Raw Sample Relinquished by: JP (WC)

Date/Time 07/25/2025 14.00
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: EM CWC