

## Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136629

Slope : 98.3

pH Meter ID : WC PH METER-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W3178         |
| BUFFER PH 7.00 GREEN 1PINT PK6 | W3093         |
| PH 10.01 BUFFER,COLOR CD 475ML | W3191         |
| buffer solution pH 7 yellow    | W3217         |
| Buffer Solution, PH2 (500ml)   | W3161         |
| pH 12.00 Buffer                | W3200         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 20.2             | 4.01        | 07/28/2025 | 08:10     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 20.2             | 7.00        | 07/28/2025 | 08:11     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 20.2             | 10.02       | 07/28/2025 | 08:14     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 20.3             | 7.00        | 07/28/2025 | 08:15     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 20.1             | 2.02        | 07/28/2025 | 09:30     |
| 6   | Q2707-03    | 1  | Water  | NA          | NA          | 24.5             | 6.10        | 07/28/2025 | 09:44     |
| 7   | Q2707-03DUP | 1  | Water  | NA          | NA          | 24.6             | 6.11        | 07/28/2025 | 09:47     |
| 8   | CCV2        | 1  | Water  | NA          | NA          | 20.3             | 12.02       | 07/28/2025 | 09:50     |

VB 136629

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph s q2708

WorkList ID : 190967

Department : Wet-Chemistry

Date : 07-28-2025 07:49:04

| Sample   | Customer Sample                                 | Matrix | Test | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method |
|----------|-------------------------------------------------|--------|------|--------------|----------|-----------------------------|--------------|--------|
| Q2707-03 | <input checked="" type="checkbox"/> OILY-LIQUID | Water  | pH   | Cool 4 deg C | PSEG03   | D41                         | 07/25/2025   | 9040C  |

Date/Time 07/28/25 08:00  
Raw Sample Received by: JFWOC  
Raw Sample Relinquished by: CD Spr

Date/Time 07/28/25 11:50  
Raw Sample Received by: CD Spr  
Raw Sample Relinquished by: JFWOC