

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB136632

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	10	28.00	82.4	82.4	07/28/2025	12:20
2	Q2707-03		1	15	100.00	>212.0	>212.0	07/28/2025	12:50
3	Q2707-03DUP		1	15	100.00	>212.0	>212.0	07/28/2025	13:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6136632

WorkList Name : fp-7-28

WorkList ID : 190975

Department : Wet-Chemistry

Date : 07-28-2025 08:18:59

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2707-03	OILY-LIQUID	Water	Flash Point	Cool 4 deg C	PSEG03	D41	07/25/2025	1010B

Date/Time 07/28/2025 12:10
Raw Sample Received by: RM CWJ
Raw Sample Relinquished by: [Signature]

Date/Time 07/28/2025 14:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]