


LB136649

7470A

INSTRUMENT ID: CV4

 μ Abs.:

40131



Linear

A= 0.0000e+000

B= 2.4450e-004

C= 8.1937e-002

Rho= 0.9994579

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.059	0.059	-92	0.000	-92	0				9
0.2	0.200	0.219	0.019	559	0.0 %	559					9
2.5	2.500	2.288	-0.212	9024	0.0 %	9024					8
5.0	5.000	5.120	0.120	20606	0.0 %	20606					2
7.5	7.500	7.619	0.119	30827	0.0 %	30827					2
10.0	10.000	9.894	-0.106	40131	0.0 %	40131					1

LB136649 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 30 Jul 2025 09:03:21

Type	Sample ID	Extended ID	Conc.	μ Abs.	Units	Type	Std Conc	Method	Date
S	0.0 - 1	50	-	-92	PPB	Std	0.0000	7470A	30 Jul 2025 09:55:43
S	0.2 - 1	50.2	-	559	PPB	Std	0.2000	7470A	30 Jul 2025 09:58:01
S	2.5 - 1	52.5	-	9024	PPB	Std	2.5000	7470A	30 Jul 2025 10:00:19
S	5.0 - 1	55.0	-	20606	PPB	Std	5.0000	7470A	30 Jul 2025 10:02:36
S	7.5 - 1	57.5	-	30827	PPB	Std	7.5000	7470A	30 Jul 2025 10:04:54
S	10.0 - 1	510	-	40131	PPB	Std	10.0000	7470A	30 Jul 2025 10:07:13
U	ICV46 - 1	ICV46	4.2280	16957	PPB	SMPL	-7470A		30 Jul 2025 10:12:36
U	ICB46 - 1	ICB46	0.0088	-299	PPB	SMPL	-7470A		30 Jul 2025 10:14:51
U	CCV68 - 1	CCV68	5.2236	21029	PPB	SMPL	-7470A		30 Jul 2025 10:17:08
U	CCB68 - 1	CCB68	0.0044	-317	PPB	SMPL	-7470A		30 Jul 2025 10:19:24
U	CRA - 1	CRA	0.2262	590	PPB	SMPL	-7470A		30 Jul 2025 10:26:53
U	HighStd - 1	HighStd	10.1677	41250	PPB	SMPL	-7470A		30 Jul 2025 10:29:07
U	ChkStd - 1	ChkStd	7.3466	29712	PPB	SMPL	-7470A		30 Jul 2025 10:31:23
U	PB169033BL - 1	PBW	-0.0393	-496	PPB	SMPL	-7470A		30 Jul 2025 10:33:42
U	PB169033BS - 1	LCSW	3.7309	14924	PPB	SMPL	-7470A		30 Jul 2025 10:36:00
U	Q2691-03 - 1	296	0.2343	623	PPB	SMPL	-7470A		30 Jul 2025 10:38:16
U	Q2691-03DUP - 1	296DUP	0.2841	827	PPB	SMPL	-7470A		30 Jul 2025 10:40:34
U	Q2691-03MS - 1	296MS	4.0365	16174	PPB	SMPL	-7470A		30 Jul 2025 10:47:40
U	Q2691-03MSD - 1	296MSD	3.5030	13992	PPB	SMPL	-7470A		30 Jul 2025 10:49:58
U	Q2691-04 - 1	297	0.0115	-288	PPB	SMPL	-7470A		30 Jul 2025 10:52:15
U	CCV69 - 1	CCV69	5.2625	21188	PPB	SMPL	-7470A		30 Jul 2025 10:54:34
U	CCB69 - 1	CCB69	0.0086	-300	PPB	SMPL	-7470A		30 Jul 2025 10:56:49
U	Q2691-05 - 1	298	31.8049	129744	PPB	SMPL	-7470A		30 Jul 2025 10:59:07
U	PB169052BL - 1	PBW	0.0863	18	PPB	SMPL	-7470A		30 Jul 2025 11:05:02
U	PB169052BS - 1	LCSW	3.5999	14388	PPB	SMPL	-7470A		30 Jul 2025 11:07:17
U	Q2703-04 - 1	TP-4	0.1497	277	PPB	SMPL	-7470A		30 Jul 2025 11:09:33
U	Q2706-02 - 1	RT-5417	0.0438	-156	PPB	SMPL	-7470A		30 Jul 2025 11:11:50
U	Q2706-04 - 1	ETGI-361	0.0844	10	PPB	SMPL	-7470A		30 Jul 2025 11:14:07
U	Q2709-02 - 1	7211935-295	0.1088	110	PPB	SMPL	-7470A		30 Jul 2025 11:16:25
U	Q2709-02DUP - 1	7211935-295DUP	0.0895	31	PPB	SMPL	-7470A		30 Jul 2025 11:18:42
U	Q2709-02MS - 1	7211935-295MS	3.6480	14585	PPB	SMPL	-7470A		30 Jul 2025 11:20:59
U	Q2709-02MSD - 1	7211935-295MSD	3.7666	15070	PPB	SMPL	-7470A		30 Jul 2025 11:23:15
U	CCV70 - 1	CCV70	5.1945	20910	PPB	SMPL	-7470A		30 Jul 2025 11:25:32
U	CCB70 - 1	CCB70	0.0274	-223	PPB	SMPL	-7470A		30 Jul 2025 11:27:49
U	PB169020TB - 1	PB169020TB	0.0810	-4	PPB	SMPL	-7470A		30 Jul 2025 11:30:07
U	Q2691-03LX5 - 1		0.0995	72	PPB	SMPL	-7470A		30 Jul 2025 11:35:20
U	Q2691-03A - 1		4.1603	16680	PPB	SMPL	-7470A		30 Jul 2025 11:37:35
U	Q2709-02LX5 - 1		0.0462	-146	PPB	SMPL	-7470A		30 Jul 2025 11:39:50
U	Q2709-02A - 1		4.1532	16651	PPB	SMPL	-7470A		30 Jul 2025 11:42:08
U	Q2691-05DLX10 - 1		2.7368	10858	PPB	SMPL	-7470A		30 Jul 2025 11:57:01
U	CCV71 - 1	CCV71	5.2390	21092	PPB	SMPL	-7470A		30 Jul 2025 11:59:18
U	CCB71 - 1	CCB71	0.0409	-168	PPB	SMPL	-7470A		30 Jul 2025 12:01:38