

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB136650

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	07/30/2025	10:00
2	CAL2	1	Water	NA	NA	20.2	7.01	07/30/2025	10:01
3	CAL3	1	Water	NA	NA	20.2	10.02	07/30/2025	10:03
4	ICV	1	Water	NA	NA	20.2	7.00	07/30/2025	10:05
5	CCV1	1	Water	NA	NA	20.3	2.02	07/30/2025	10:10
6	Q2719-01	1	Water	NA	NA	23.6	5.70	07/30/2025	10:20
7	Q2719-01DUP	1	Water	NA	NA	23.8	5.72	07/30/2025	10:23
8	Q2719-02	1	Water	NA	NA	22.3	8.45	07/30/2025	10:30
9	CCV2	1	Water	NA	NA	20.3	12.02	07/30/2025	10:32

WORKLIST(Hardcopy Internal Chain)

136650

WorkList Name : ph w q2719 WorkList ID : 191011 Department : Wet-Chemistry Date : 07-30-2025 09:33:41

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2719-01	A 50738	Water	pH	Cool 4 deg C	PSEG03	D31	07/29/2025	9040C
Q2719-02	C 50765	Water	pH	Cool 4 deg C	PSEG03	D31	07/29/2025	9040C

Date/Time 07/30/25 09:16:55
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 07/30/25 12:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]