

Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB136660
Analysis Date: 07/31/2025
BalanceID: WC-SC-6
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: Iwona
Extraction Date: 07/31/2025
Extraction IN Time: 08:40
Extraction OUT Time: 09:35
Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB136660BL	LB136660BL	WATER	1.3	1000	100	3.2853	3.2853	3.02	3.2854	3.2854	0.0001	0.1
2	LB136660BS	LB136660BS	WATER	1.3	1000	100	3.1403	3.1403	3.01	3.1573	3.1573	0.0170	17
3	LB136660BSD	LB136660BSD	WATER	1.3	1000	100	2.8503	2.8503	3.03	2.8675	2.8675	0.0172	17.2
4	Q2702-01	MH-7-25-25	WATER	1.6	1000	100	3.0227	3.0227	3.04	3.0544	3.0544	0.0317	31.7
5	Q2725-01	Grab	WATER	1.6	1000	100	3.0680	3.0680	3.05	3.0844	3.0844	0.0164	16.4

QC Batch# LB136660

Test: TPH

Analysis Date: 07/31/2025

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W3204
pH Paper 0-14	M6069
Sodium Sulfate	EP2629
1:1 HCL	WP112782
Silica Gel	W3079
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP112783
LCSWD	5.00 ML	WP112784
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-6

Before Analysis

0.0020 gram Balance: 0.0018 (0.0018-0.0022) In OVEN TEMP1 : 70 °C Dessicator Time In1 : 11:26

1.0000 gram Balance: 1.0004 (0.9950-1.0050) In Time1: 10:30

Bal Check Time: 09:00 Out OVEN TEMP1: 70 °C Dessicator Time Out1: 12:00

Out Time1: 11:25

After Analysis

0.0020 gram Balance: 0.0019 (0.0018-0.0022) In OVEN TEMP2 : 71 °C Dessicator Time In2 : 13:01

1.0000 gram Balance: 1.0005 (0.9950-1.0050) In Time2: 12:30

Bal Check Time: 13:35 Out OVEN TEMP2: 71 °C Dessicator Time Out2: 13:30

Out Time2: 13:00

WORKLIST(Hardcopy Internal Chain)

13660

WorkList Name : TPH Q2730 WorkList ID : 191031 Department : Wet-Chemistry Date : 07-31-2025 07:50:55

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2702-01	MH-7-25-25	Water	TPH	Conc H2SO4 to pH < 2	EURO03	D31	07/25/2025	1664A
Q2725-01	Grab	Water	TPH	Conc H2SO4 to pH < 2	ARAM01	O22	07/30/2025	1664A

Date/Time 07/31/25 08:10
 Raw Sample Received by: SP CSM
 Raw Sample Relinquished by: SP CSM

Date/Time 07/31/25 19:30
 Raw Sample Received by: SP CSM
 Raw Sample Relinquished by: SP CSM