

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Eman

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB136663

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.7	07/31/2025	09:30
2	Q2719-01		1	12	100.00	>212.0	>212.0	07/31/2025	10:00
3	Q2719-01DUP		1	12	100.00	>212.0	>212.0	07/31/2025	10:30
4	Q2719-02		1	14	100.00	>212.0	>212.0	07/31/2025	11:00
5	Q2733-01		1	15	100.00	>212.0	>212.0	07/31/2025	11:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16136663

WorkList Name : FP-07-29

WorkList ID : 191006

Department : Wet-Chemistry

Date : 07-29-2025 12:06:47

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2719-01	50738	Water	Flash Point	Cool 4 deg C	PSEG03	D31	07/29/2025	1010B
Q2719-02	50765	Water	Flash Point	Cool 4 deg C	PSEG03	D31	07/29/2025	1010B
Q2733-01	TW-WTS-12	Water	Flash Point	Cool 4 deg C	ENTA05	O11	07/30/2025	1010B

Date/Time 07/31/25 09:00
 Raw Sample Received by: EM(wc)
 Raw Sample Relinquished by: EM(wc)

Date/Time 07/31/25 12:15
 Raw Sample Received by: EM(wc)
 Raw Sample Relinquished by: EM(wc)