

Analytical Summary Report

Analysis Method: 9045D

Analyst By : Eman

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB136676

Slope : 99.4

BalanceID: WC SC-7

pH Meter ID : WC pH meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	08/01/2025	11:00
2	CAL2	1	Water	NA	NA	20.3	7.01	08/01/2025	11:03
3	CAL3	1	Water	NA	NA	20.3	9.99	08/01/2025	11:05
4	ICV	1	Water	NA	NA	20.3	7.00	08/01/2025	11:08
5	CCV1	1	Water	NA	NA	20.4	2.00	08/01/2025	12:10
6	Q2734-02	1	Solid	20	20	22.4	8.61	08/01/2025	12:13
7	Q2734-02DUP	1	Solid	20	20	22.4	8.60	08/01/2025	12:16
8	Q2737-01	1	Solid	20	20	22.6	5.47	08/01/2025	12:20
9	Q2743-02	1	Solid	20	20	24.4	7.75	08/01/2025	12:24
10	CCV2	1	Water	NA	NA	20.4	12.02	08/01/2025	12:30

WORKLIST(Hardcopy Internal Chain)

LB136676

WorkList Name : Corrosivity-080125

WorkList ID : 191061

Department : Wet-Chemistry

Date : 08-01-2025 09:07:10

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2734-02	VNJ-253	Solid	Corrosivity	Cool 4 deg C	PSEG03	D31	07/31/2025	9045D
Q2737-01	OILY-DEBRIS	Solid	Corrosivity	Cool 4 deg C	PSEG03	D21	07/31/2025	9045D
Q2743-02	WC1	Solid	Corrosivity	Cool 4 deg C	GENV01	O13	07/31/2025	9045D

Date/Time 08/01/25 10:50
 Raw Sample Received by: EM(wbc)
 Raw Sample Relinquished by: AP SM

Date/Time 08/01/25 12:40
 Raw Sample Received by: AP SM
 Raw Sample Relinquished by: EM(wbc)