

Analysis Method: 9034

Parameter: Reactive Sulfide

Run Number: LB136677

ANALYST: rubina

SUPERVISOR REVIEW BY: Iwona

Constant: 16000

Normality1: 0.025

Normality2: 0.025

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE	W3105
IODINE SOLUTION .025N 1L	W3213
Starch Solution, 4L	W3149

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB169084BL		1	50	50	2.00	0.00	1.94	1.94	0.06	0.00	0.00	08/01/2025	14:00
2	Q2736-02		1	50	50	2.00	0.00	1.90	1.90	0.10	0.04	0.32	08/01/2025	14:03
3	Q2736-02DUP		1	50	50	2.00	0.00	1.90	1.90	0.10	0.04	0.32	08/01/2025	14:06
4	Q2736-03		1	50	50	2.00	0.00	1.86	1.86	0.14	0.08	0.64	08/01/2025	14:09
5	Q2736-04		1	50	50	2.00	0.00	1.86	1.86	0.14	0.08	0.64	08/01/2025	14:12
6	Q2737-02		1	50	50	2.00	0.00	1.90	1.90	0.10	0.04	0.32	08/01/2025	14:15

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume