

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Eman

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB136679

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	10	28.00	82.4	82.1	08/01/2025	13:00
2	Q2736-02		1	14	100.00	>212.0	>212.0	08/01/2025	13:30
3	Q2736-02DUP		1	14	100.00	>212.0	>212.0	08/01/2025	14:00
4	Q2736-03		1	12	100.00	>212.0	>212.0	08/01/2025	14:30
5	Q2736-04		1	15	100.00	>212.0	>212.0	08/01/2025	15:00
6	Q2737-02		1	13	100.00	>212.0	>212.0	08/01/2025	15:30
7	Q2739-01		1	15	100.00	>212.0	>212.0	08/01/2025	16:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

16136679

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp-8-01

WorkList ID : 191054

Department : Wet-Chemistry

Date : 08-01-2025 08:26:39

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2736-02	4023	Water	Flash Point	Cool 4 deg C	PSEG03	D21	07/31/2025	1010B
Q2736-03	4043	Water	Flash Point	Cool 4 deg C	PSEG03	D21	07/31/2025	1010B
Q2736-04	4044	Water	Flash Point	Cool 4 deg C	PSEG03	D21	07/31/2025	1010B
Q2737-02	OILY-WATER	Water	Flash Point	Cool 4 deg C	PSEG03	D21	07/31/2025	1010B
Q2739-01	50743	Water	Flash Point	Cool 4 deg C	PSEG03	D21	07/31/2025	1010B

Date/Time 08/01/25 12:50
Raw Sample Received by: EM(WC)
Raw Sample Relinquished by: CP SM

Date/Time 08/01/25 16:40
Raw Sample Received by: CP SM
Raw Sample Relinquished by: EM(WC)