

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136683

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	08/04/2025	08:15
2	CAL2	1	Water	NA	NA	20.2	7.00	08/04/2025	08:16
3	CAL3	1	Water	NA	NA	20.3	10.02	08/04/2025	08:17
4	ICV	1	Water	NA	NA	20.3	7.00	08/04/2025	08:20
5	CCV1	1	Water	NA	NA	20.2	2.01	08/04/2025	08:25
6	Q2751-01	1	Water	NA	NA	20.4	6.89	08/04/2025	08:35
7	Q2751-01DUP	1	Water	NA	NA	20.6	6.90	08/04/2025	08:37
8	Q2754-02	1	Water	NA	NA	20.1	8.90	08/04/2025	08:45
9	Q2754-02DUP	1	Water	NA	NA	20.3	8.92	08/04/2025	08:49
10	Q2756-01	1	Water	NA	NA	20.2	6.45	08/04/2025	08:55
11	Q2756-01DUP	1	Water	NA	NA	20.3	6.47	08/04/2025	08:59
12	CCV2	1	Water	NA	NA	20.3	12.02	08/04/2025	09:00

WORKLIST(Hardcopy Internal Chain)

20136683

WorkList Name : ph q2756

WorkList ID : 191071

Department : Wet-Chemistry

Date : 08-04-2025 08:09:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2751-01	ROLL-OFF-LIQUID	Water	pH	Cool 4 deg C	PSEG03	D31	08/01/2025	9040C
Q2754-02	OILY-WATER	Water	pH	Cool 4 deg C	PSEG03	D41	08/01/2025	9040C
Q2756-01	71725	Water	pH	Cool 4 deg C	PSEG03	D41	08/01/2025	9040C

Date/Time 08/04/25 08:10

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 08/04/25 12:30

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]