

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB136690

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	08/04/2025	08:25
2	Q2751-01		1	14	100.00	>212.0	>212.0	08/04/2025	08:55
3	Q2751-01DUP		1	14	100.00	>212.0	>212.0	08/04/2025	09:25
4	Q2754-02		1	13	100.00	>212.0	>212.0	08/04/2025	09:55
5	Q2756-01		1	15	100.00	>212.0	>212.0	08/04/2025	10:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16136698

WorkList Name : flash point-8-04 WorkList ID : 191070 Department : Wet-Chemistry Date : 08-04-2025 08:09:24

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2751-01	ROLL-OFF-LIQUID	Water	Flash Point	Cool 4 deg C	PSEG03	D31	08/01/2025	1010B
Q2754-02	OILY-WATER	Water	Flash Point	Cool 4 deg C	PSEG03	D41	08/01/2025	1010B
Q2756-01	71725	Water	Flash Point	Cool 4 deg C	PSEG03	D41	08/01/2025	1010B

Date/Time 08/04/2025 08.15
 Raw Sample Received by: RM (wg)
 Raw Sample Relinquished by: JSC (wg)

Date/Time 08/04/2025 11.00
 Raw Sample Received by: JSC (wg)
 Raw Sample Relinquished by: RM (wg)