

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB136691

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	08/04/2025	11:10
2	Q2722-02		1	13	100.00	>212.0	>212.0	08/04/2025	11:40
3	Q2722-02DUP		1	13	100.00	>212.0	>212.0	08/04/2025	12:10
4	Q2722-04		1	15	100.00	>212.0	>212.0	08/04/2025	12:40
5	Q2752-03		1	11	61.00	141.8	141.5	08/04/2025	13:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB136691

WorkList Name : fp-8-4

WorkList ID : 191069

Department : Wet-Chemistry

Date : 08-04-2025 08:08:12

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2722-02	LAW-25-0107	Solid	Flash Point	Cool 4 deg C	PSEG03	D21	07/29/2025	1010B
Q2722-04	LAW-25-0108	Solid	Flash Point	Cool 4 deg C	PSEG03	D21	07/29/2025	1010B
Q2752-03	PIPE-LINE-LIQUIDS	Solid	Flash Point	Cool 4 deg C	PSEG03	D21	08/01/2025	1010B

Date/Time 08/04/2025 08:15
Raw Sample Received by: RM CWO
Raw Sample Relinquished by: RM CWO

Date/Time 08/04/2025 13:50
Raw Sample Received by: RM CWO
Raw Sample Relinquished by: RM CWO