

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB136696

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	08/04/2025	15:00
2	CAL2	1	Water	NA	NA	20.2	7.00	08/04/2025	15:01
3	CAL3	1	Water	NA	NA	20.2	10.02	08/04/2025	15:03
4	ICV	1	Water	NA	NA	20.3	7.02	08/04/2025	15:10
5	CCV1	1	Water	NA	NA	20.2	2.01	08/04/2025	15:11
6	Q2759-04	1	Water	NA	NA	20.4	6.15	08/04/2025	15:30
7	Q2759-04DUP	1	Water	NA	NA	20.6	6.17	08/04/2025	15:33
8	CCV2	1	Water	NA	NA	20.2	12.02	08/04/2025	15:35

WORKLIST(Hardcopy Internal Chain)

VB 136696

WorkList Name : ph w q2759 WorkList ID : 191097 Department : Wet-Chemistry Date : 08-04-2025 14:31:10

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2759-04	K LIQUID-DRUMS	Water	pH	Cool 4 deg C	PSEG03	D31	08/04/2025	9040C

Date/Time 08/04/25 16:45
 Raw Sample Received by: SP CAGY
 Raw Sample Relinquished by: L. C. Sma

Date/Time 08/04/25 17:30
 Raw Sample Received by: L. C. Sma
 Raw Sample Relinquished by: SP CAGY