

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB136700

Slope : 98.3

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.02	08/05/2025	08:45
2	CAL2	1	Water	NA	NA	20.2	7.00	08/05/2025	08:46
3	CAL3	1	Water	NA	NA	20.3	10.02	08/05/2025	08:47
4	ICV	1	Water	NA	NA	20.3	6.99	08/05/2025	08:50
5	CCV1	1	Water	NA	NA	20.2	2.01	08/05/2025	08:52
6	Q2758-02	1	Solid	20.02	20	20.4	7.37	08/05/2025	09:10
7	Q2759-02	1	Solid	20.03	20	20.6	6.90	08/05/2025	09:20
8	Q2763-04	1	Solid	20.04	20	22.7	8.50	08/05/2025	09:30
9	Q2763-04DUP	1	Solid	20.05	20	22.9	8.52	08/05/2025	09:34
10	CCV2	1	Water	NA	NA	20.3	12.02	08/05/2025	09:35

WORKLIST(Hardcopy Internal Chain)

136700

WorkList Name : corrosivity q2763 WorkList ID : 191106 Department : Wet-Chemistry Date : 08-05-2025 08:26:02

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2758-02	WASTE	Solid	Corrosivity	Cool 4 deg C	SCIA01	D21	08/01/2025	9045D
Q2759-02	SOLID-DRUMS	Solid	Corrosivity	Cool 4 deg C	PSEG03	D31	08/04/2025	9045D
Q2763-04	TP-2	Solid	Corrosivity	Cool 4 deg C	PSEG03	O23	08/04/2025	9045D

Date/Time 08/05/25 08:35
Raw Sample Received by: 18 WDC
Raw Sample Relinquished by: RJ C ERT-146)

Date/Time 08/05/25 12:30
Raw Sample Received by: RJ C ERT-146)
Raw Sample Relinquished by: 38 GDC