

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB136711

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	10	28.00	82.4	81.7	08/05/2025	14:30
2	Q2759-04		1	15	100.00	>212.0	>212.0	08/05/2025	15:00
3	Q2776-01		1	14	100.00	>212.0	>212.0	08/05/2025	15:30
4	Q2776-01DUP		1	14	100.00	>212.0	>212.0	08/05/2025	16:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

6136711

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp-8-05 WorkList ID : 191110 Department : Wet-Chemistry Date : 08-05-2025 09:36:02

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2759-04	LIQUID-DRUMS	Water	Flash Point	Cool 4 deg C	PSEG03	D31	08/04/2025	1010B
Q2776-01	72825	Water	Flash Point	Cool 4 deg C	PSEG03	D31	08/05/2025	1010B

Date/Time 08/05/2025 14:05
Raw Sample Received by: RM CW12
Raw Sample Relinquished by: RM CW12

Date/Time 08/05/2025 16:40
Raw Sample Received by: RM CW12
Raw Sample Relinquished by: RM CW12