

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136715

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	08/06/2025	08:30
2	CAL2	1	Water	NA	NA	20.2	7.00	08/06/2025	08:31
3	CAL3	1	Water	NA	NA	20.3	10.02	08/06/2025	08:35
4	ICV	1	Water	NA	NA	20.2	7.02	08/06/2025	08:37
5	CCV1	1	Water	NA	NA	20.2	2.01	08/06/2025	08:40
6	Q2776-01	1	Water	NA	NA	22.3	5.51	08/06/2025	08:50
7	Q2776-01DUP	1	Water	NA	NA	22.5	5.52	08/06/2025	08:52
8	CCV2	1	Water	NA	NA	20.3	12.02	08/06/2025	09:00

WORKLIST(Hardcopy Internal Chain)

136715

WorkList Name : ph s q2776 WorkList ID : 191123 Department : Wet-Chemistry Date : 08-06-2025 08:17:48

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2776-01	A 72825	Water	pH	Cool 4 deg C	PSEG03	D31	08/05/2025	9040C

Date/Time 08/06/25 08:25
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 08/06/25 11:50
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]