

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136716

Slope : 98.2

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	08/06/2025	08:30
2	CAL2	1	Water	NA	NA	20.2	7.00	08/06/2025	08:31
3	CAL3	1	Water	NA	NA	20.3	10.02	08/06/2025	08:35
4	ICV	1	Water	NA	NA	20.2	7.02	08/06/2025	08:37
5	CCV1	1	Water	NA	NA	20.2	2.02	08/06/2025	09:10
6	Q2775-01	1	Solid	20.02	20	21.9	7.27	08/06/2025	09:20
7	Q2775-01DUP	1	Solid	20.04	20	21.8	7.29	08/06/2025	09:23
8	CCV2	1	Water	NA	NA	20.2	12.02	08/06/2025	09:25

WORKLIST(Hardcopy Internal Chain)

136716

WorkList Name : ph q2775 WorkList ID : 191122 Department : Wet-Chemistry Date : 08-06-2025 08:17:01

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2775-01	F	60273	Solid	pH	Cool 4 deg C	PSEG03 D31	08/05/2025	9045D

Date/Time 08/06/25 08:25
 Raw Sample Received by: ef wpc/
 Raw Sample Relinquished by: Al sm

Date/Time 08/06/25 11:50
 Raw Sample Received by: Al sm
 Raw Sample Relinquished by: ef wpc/