


LB136732
INSTRUMENT ID: CV1

7470A

μ Abs.:

36724



Linear

A= 0.0000e+000

B= 2.7036e-004

C= -1.8142e-002

Rho= 0.9997683

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.004	-0.004	54	0.000	54	0			
0.2	0.200	0.179	-0.021	730	0.0 %	730				
2.5	2.500	2.535	0.035	9442	0.0 %	9442	0			
5.0	5.000	4.928	-0.072	18294	0.0 %	18294				
7.5	7.500	7.652	0.152	28369	0.0 %	28369				
10.0	10.000	9.910	-0.090	36724	0.0 %	36724				

LB136732 INSTRUMENT ID: CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 Aug 2025 09:14:39

Type	Sample ID	Extended ID	Conc	µAbs	Units	Type	Std Conc	Method	Date
S	0.0 - 1	50.0	-	54	PPB	Std	0.000	7470A	07 Aug 2025 09:27:08
S	0.2 - 1	50.2	-	73	PPB	Std	0.200	7470A	07 Aug 2025 09:29:25
S	2.5 - 1	52.5	-	944	PPB	Std	2.500	7470A	07 Aug 2025 09:31:42
S	5.0 - 1	55.0	-	1829	PPB	Std	5.000	7470A	07 Aug 2025 09:33:59
S	7.5 - 1	57.5	-	2836	PPB	Std	7.500	7470A	07 Aug 2025 09:36:17
S	10.0 - 1	510.0	-	3672	PPB	Std	10.000	7470A	07 Aug 2025 09:38:35
U	ICV56 - 1	ICV56	4.0285	1493	PPB	SMPL	7470A	7470A	07 Aug 2025 09:41:49
U	ICB56 - 1	ICB56	-0.0638	-16	PPB	SMPL	7470A	7470A	07 Aug 2025 09:46:34
U	CCV95 - 1	CCV95	4.8842	1813	PPB	SMPL	7470A	7470A	07 Aug 2025 09:48:49
U	CCB95 - 1	CCB95	-0.0725	-201	PPB	SMPL	7470A	7470A	07 Aug 2025 09:51:05
U	CRA - 1	CRA	0.1622	66	PPB	SMPL	7470A	7470A	07 Aug 2025 09:58:28
U	HighStd - 1	HighStd	9.9366	36821	PPB	SMPL	7470A	7470A	07 Aug 2025 10:00:43
U	ChkStd - 1	ChkStd	7.4253	2753	PPB	SMPL	7470A	7470A	07 Aug 2025 10:02:58
U	PB169158BL - 1	PBW	-0.0609	-15	PPB	SMPL	7470A	7470A	07 Aug 2025 10:07:53
U	PB169158BS - 1	LCSW	4.2416	1575	PPB	SMPL	7470A	7470A	07 Aug 2025 10:10:09
U	2747-02 - 1	OU4-TS-GRILLO-OG2-073125	-0.0930	-27	PPB	SMPL	7470A	7470A	07 Aug 2025 10:12:25
U	2747-04 - 1	OU4-TS-GRILLO-OG3-073125	0.0246	1	PPB	SMPL	7470A	7470A	07 Aug 2025 10:14:43
U	2747-06 - 1	OU4-TS-GRILLO-OG4-073125	0.1744	71	PPB	SMPL	7470A	7470A	07 Aug 2025 10:17:01
U	2747-08 - 1	OU4-TS-GRILLO-TSCP01-073125	-0.0246	-2	PPB	SMPL	7470A	7470A	07 Aug 2025 10:19:18
U	2747-10 - 1	OU4-TS-GRILLO-TSCP02-073125	0.0695	32	PPB	SMPL	7470A	7470A	07 Aug 2025 10:21:35
U	CCV96 - 1	CCV96	4.9759	1847	PPB	SMPL	7470A	7470A	07 Aug 2025 10:23:50
U	CCB96 - 1	CCB96	-0.0592	-15	PPB	SMPL	7470A	7470A	07 Aug 2025 10:26:05
U	2747-10DUP - 1	OU4-TS-GRILLO-TSCP02-073125DUP	0.0362	201	PPB	SMPL	7470A	7470A	07 Aug 2025 10:28:22
U	2747-10MS - 1	OU4-TS-GRILLO-TSCP02-073125MS	3.9264	1459	PPB	SMPL	7470A	7470A	07 Aug 2025 10:30:38
U	2747-10MSD - 1	OU4-TS-GRILLO-TSCP02-073125MSD	3.9561	1470	PPB	SMPL	7470A	7470A	07 Aug 2025 10:32:53
U	PB169102TB - 1	PB169102TB PB169158TB	-0.0695	-1	PPB	SMPL	7470A	7470A	07 Aug 2025 10:35:11
U	2747-10LX5 - 1		-0.0419	-88	PPB	SMPL	7470A	7470A	07 Aug 2025 10:37:28
U	2747-10A - 1		4.0696	1512	PPB	SMPL	7470A	7470A	07 Aug 2025 10:39:45
U	CV97 - 1	CCV97	4.8166	1788	PPB	SMPL	7470A	7470A	07 Aug 2025 10:42:01
U	CCB97 - 1	CCB97	-0.0192	-4	PPB	SMPL	7470A	7470A	07 Aug 2025 10:44:20