

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : rubina

Run Number: LB136796

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	08/13/2025	08:00
2	CAL2	1	Water	NA	NA	20.2	7.00	08/13/2025	08:01
3	CAL3	1	Water	NA	NA	20.2	10.02	08/13/2025	08:02
4	ICV	1	Water	NA	NA	20.3	7.02	08/13/2025	08:05
5	CCV1	1	Water	NA	NA	20.3	2.01	08/13/2025	08:10
6	Q2840-03	1	Water	NA	NA	25.6	6.61	08/13/2025	08:20
7	Q2840-03DUP	1	Water	NA	NA	25.9	6.63	08/13/2025	08:22
8	CCV2	1	Water	NA	NA	20.3	12.02	08/13/2025	08:25

VB 136796

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph w q2840 WorkList ID : 191238 Department : Wet-Chemistry Date : 08-13-2025 07:34:02

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2840-03 C	0804-E	Water	pH	Cool 4 deg C	PSEG03	D31	08/12/2025	9040C

Date/Time 08/13/25 07:40
Raw Sample Received by: WEC
Raw Sample Relinquished by: CP SM

Date/Time 08/13/25 10:40
Raw Sample Received by: CP SM
Raw Sample Relinquished by: WEC