

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : rubina

Run Number: LB136814

Slope : 99.2

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	08/14/2025	09:00
2	CAL2	1	Water	NA	NA	20.2	6.99	08/14/2025	09:01
3	CAL3	1	Water	NA	NA	20.3	10.02	08/14/2025	09:05
4	ICV	1	Water	NA	NA	20.3	7.01	08/14/2025	09:10
5	CCV1	1	Water	NA	NA	20.3	2.01	08/14/2025	09:15
6	Q2847-01	1	Solid	20.02	20	23.0	6.53	08/14/2025	09:30
7	Q2847-01DUP	1	Solid	20.03	20	23.2	6.55	08/14/2025	09:34
8	Q2847-02	1	Solid	20.05	20	22.8	6.81	08/14/2025	09:37
9	CCV2	1	Water	NA	NA	20.3	12.02	08/14/2025	09:40

WORKLIST(Hardcopy Internal Chain)

136814

WorkList Name : PH S Q2847

WorkList ID : 191258

Department : Wet-Chemistry

Date : 08-14-2025 08:12:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2847-01	BUR-25-0069	Solid	pH	Cool 4 deg C	PSEG03	J41	08/13/2025	9045D
Q2847-02	BUR-25-0070	Solid	pH	Cool 4 deg C	PSEG03	J41	08/13/2025	9045D

Date/Time 08-14-25 08:20

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 08-14-25

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]